

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: AMANDEEP

ANALYST: Aparna

Date: 09/27/2018

TEMP1 IN: 103 09/27/2018 07:00 TEMP OUT: 104 09/27/2018 08:05

TEMP2 IN: 104 09/27/2018 09:10 TEMP OUT: 104 09/27/2018 10:20

TEMP3 IN: 105 09/27/2018 11:30 TEMP OUT: 105 09/27/2018 13:05

TEMP4 IN: 105 09/27/2018 13:45 TEMP OUT: 105 09/27/2018 15:20

Run Number: LB98172

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB98172BL	LB98172BL	1.4025	1.4025	100	1.4025	1.4025	1.4025	0.0000	0
2	LB98172BS	LB98172BS	1.3928	1.3928	100	1.4430	1.4430	1.4430	0.0502	502
3	J5076-01	092418-SIDI-F-D-B	1.4037	1.4037	200	1.4233	1.4233	1.4233	0.0196	98
4	J5076-01DUP	092418-SIDI-F-D-BDUP	1.4053	1.4053	200	1.4250	1.4250	1.4250	0.0197	98.5
5	J5076-02	092418-SIDI-F-D-M	1.4032	1.4032	400	1.4195	1.4195	1.4195	0.0163	40.7
6	J5076-03	092418-SIDI-F-D-S	1.4070	1.4070	400	1.4132	1.4132	1.4132	0.0062	15.5
7	J5076-04	092418-SIDI-F-U-B	1.3872	1.3872	200	1.4085	1.4085	1.4085	0.0213	106.5
8	J5076-05	092418-SIDI-F-U-M	1.3963	1.3963	400	1.4107	1.4107	1.4107	0.0144	36
9	J5076-06	092418-SIDI-F-U-S	1.4085	1.4085	400	1.4151	1.4151	1.4151	0.0066	16.5
10	J5076-07	092418-TIPC-F-D-B	1.3843	1.3843	200	1.4089	1.4089	1.4089	0.0246	123
11	J5076-08	092418-TIPC-F-D-S	1.4068	1.4068	400	1.4190	1.4190	1.4190	0.0122	30.5
12	J5076-09	092418-TIPC-F-U-B	1.3918	1.3918	500	1.4036	1.4036	1.4036	0.0118	23.6
13	J5076-10	092418-TIPC-F-U-S	1.3963	1.3963	400	1.4062	1.4062	1.4062	0.0099	24.8
14	J5076-11	092418-DBRI-E-D-B	1.4041	1.4041	500	1.4185	1.4185	1.4185	0.0144	28.8
15	J5076-12	092418-DBRI-E-D-S	1.3877	1.3877	400	1.3948	1.3948	1.3948	0.0071	17.8



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: AMANDEEP

ANALYST: Aparna

Date: 09/27/2018

TEMP1 IN: 103 09/27/2018 07:00 TEMP OUT: 104 09/27/2018 08:05

TEMP2 IN: 104 09/27/2018 09:10 TEMP OUT: 104 09/27/2018 10:20

TEMP3 IN: 105 09/27/2018 11:30 TEMP OUT: 105 09/27/2018 13:05

TEMP4 IN: 105 09/27/2018 13:45 TEMP OUT: 105 09/27/2018 15:20

Run Number: LB98172

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J5076-13	092418-DBRI-E-U-B	1.4017	1.4017	400	1.4123	1.4123	1.4123	0.0106	26.5
17	J5076-14	092418-DBRI-E-U-S	1.3991	1.3991	200	1.4048	1.4048	1.4048	0.0057	28.5
18	J5076-14DUP	092418-DBRI-E-U-SDUP	1.3934	1.3934	200	1.3990	1.3990	1.3990	0.0056	28

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS092718

WorkList ID : 117227

Date : 9/27/2018 7:22:18 AM

2098172

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/01/2018	Water	J5076-01	TSS	Cool 4 deg C	HDR103	M43	092418-SIDI-F-D-B	09/24/2018	SM2540 D
10/01/2018	Water	J5076-02	TSS	Cool 4 deg C	HDR103	M43	092418-SIDI-F-D-M	09/24/2018	SM2540 D
10/01/2018	Water	J5076-03	TSS	Cool 4 deg C	HDR103	M43	092418-SIDI-F-D-S	09/24/2018	SM2540 D
10/01/2018	Water	J5076-04	TSS	Cool 4 deg C	HDR103	M43	092418-SIDI-F-U-B	09/24/2018	SM2540 D
10/01/2018	Water	J5076-05	TSS	Cool 4 deg C	HDR103	M43	092418-SIDI-F-U-M	09/24/2018	SM2540 D
10/01/2018	Water	J5076-06	TSS	Cool 4 deg C	HDR103	M43	092418-SIDI-F-U-S	09/24/2018	SM2540 D
09/27/2018	Water	J5076-07	TSS	Cool 4 deg C	HDR103	M43	092418-TIPC-F-D-B	09/24/2018	SM2540 D
09/27/2018	Water	J5076-08	TSS	Cool 4 deg C	HDR103	M43	092418-TIPC-F-D-S	09/24/2018	SM2540 D
09/27/2018	Water	J5076-09	TSS	Cool 4 deg C	HDR103	M43	092418-TIPC-F-U-B	09/24/2018	SM2540 D
09/27/2018	Water	J5076-10	TSS	Cool 4 deg C	HDR103	M43	092418-TIPC-F-U-S	09/24/2018	SM2540 D
10/01/2018	Water	J5076-11	TSS	Cool 4 deg C	HDR103	M43	092418-DBRI-E-D-B	09/24/2018	SM2540 D
10/01/2018	Water	J5076-12	TSS	Cool 4 deg C	HDR103	M43	092418-DBRI-E-D-S	09/24/2018	SM2540 D
10/01/2018	Water	J5076-13	TSS	Cool 4 deg C	HDR103	M43	092418-DBRI-E-U-B	09/24/2018	SM2540 D
10/01/2018	Water	J5076-14	TSS	Cool 4 deg C	HDR103	M43	092418-DBRI-E-U-S	09/24/2018	SM2540 D

fp-9823888

Date/Time 09/27/18, 08:00
 Received by: AS
 Relinquished by: [Signature]

Date/Time 09/27/18, 12:30
 Received by: [Signature]
 Relinquished by: AS