



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB98183

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.6	09/27/2018	09:40
2	J5083-04		1	7	>100.0	>212.0	>212.0	09/27/2018	10:15
3	J5083-04DUP		1	7	>100.0	>212.0	>212.0	09/27/2018	10:40
4	J5083-05		1	9	>100.0	>212.0	>212.0	09/27/2018	11:10
5	J5083-06		1	7	>100.0	>212.0	>212.0	09/27/2018	11:40
6	J5085-08		1	6	>100.0	>212.0	>212.0	09/27/2018	12:05
7	J5085-09		1	8	>100.0	>212.0	>212.0	09/27/2018	12:45
8	J5085-10		1	6	>100.0	>212.0	>212.0	09/27/2018	13:10
9	J5085-11		1	9	>100.0	>212.0	>212.0	09/27/2018	13:42
10	J5085-12		1	7	>100.0	>212.0	>212.0	09/27/2018	14:15
11	J5085-13		1	6	>100.0	>212.0	>212.0	09/27/2018	14:48
12	J5085-14		1	8	>100.0	>212.0	>212.0	09/27/2018	15:20
13	J5090-01		1	5	>100.0	>212.0	>212.0	09/27/2018	15:50
14	J5090-02		1	7	>100.0	>212.0	>212.0	09/27/2018	16:10
15	J5090-03		1	8	>100.0	>212.0	>212.0	09/27/2018	16:40
16	J5101-02		1	7	>100.0	>212.0	>212.0	09/27/2018	17:02

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp92718

WorkList ID : 117258

Date : 9/27/2018 9:04:07 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2018	Water	J5083-04 ✓	Flash Point	Cool 4 deg C	ENV177	M63	TKA-092518	09/25/2018	1010A
10/02/2018	Water	J5083-05 ✓	Flash Point	Cool 4 deg C	ENV177	M63	TKC-092518	09/25/2018	1010A
10/02/2018	Water	J5083-06 ✓	Flash Point	Cool 4 deg C	ENV177	M63	TKB-092518	09/25/2018	1010A
10/08/2018	Water	J5085-08 B	Flash Point	Cool 4 deg C	TULL02	M61	MH-720	09/25/2018	1010A
10/08/2018	Water	J5085-09 B	Flash Point	Cool 4 deg C	TULL02	M61	MH-719	09/25/2018	1010A
10/08/2018	Water	J5085-10 B	Flash Point	Cool 4 deg C	TULL02	M61	MH-717	09/25/2018	1010A
10/08/2018	Water	J5085-11 B	Flash Point	Cool 4 deg C	TULL02	M61	MH-10.25	09/25/2018	1010A
10/08/2018	Water	J5085-12 B	Flash Point	Cool 4 deg C	TULL02	M61	MH-10.29	09/25/2018	1010A
10/08/2018	Water	J5085-13 B	Flash Point	Cool 4 deg C	TULL02	M61	MH-714	09/25/2018	1010A
10/08/2018	Water	J5085-14 B	Flash Point	Cool 4 deg C	TULL02	M61	MH-713	09/25/2018	1010A
10/02/2018	Water	J5090-01 H	Flash Point	Cool 4 deg C	TULL02	M63	MW-1	09/25/2018	1010A
10/02/2018	Water	J5090-02 H	Flash Point	Cool 4 deg C	TULL02	M63	MW-2	09/25/2018	1010A
10/02/2018	Water	J5090-03 H	Flash Point	Cool 4 deg C	TULL02	M63	MW-4	09/25/2018	1010A
10/01/2018	Water	J5101-02 C	Flash Point	Cool 4 deg C	PSEG01	I33	C-1329	09/26/2018	1010A

Date/Time 09/27/18 9:10 AM
 Received by: KS
 Relinquished by: CP

Date/Time 09/27/18 17:05
 Received by: CP
 Relinquished by: KS