

LB 98188


7471B

INSTRUMENT ID: CY1

Linear

 μ Abs.:

42559



A= 0.0000e+000

B= 2.3368e-004

C= -4.1802e-002

Rho= 0.9996380

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.040	-0.040	7	0.000	7				
0.2	0.200	0.197	-0.003	1024	0.0 %	1024				
2.5	2.500	2.425	-0.075	10554	0.0 %	10554				
5.0	5.000	5.206	0.206	22457	0.0 %	22457				
7.5	7.500	7.509	0.009	32310	0.0 %	32310				
10.0	10.000	9.904	-0.096	42559	0.0 %	42559				

LB98188 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Sep 2018 14:36:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	7	-	0.00007471B	PPB		27 Sep 2018 14:41:17	S	Std
0.2 - 1	50.2	1024	-	0.20007471B	PPB		27 Sep 2018 14:43:34	S	Std
2.5 - 1	52.5	10554	-	2.50007471B	PPB		27 Sep 2018 14:45:52	S	Std
5.0 - 1	55	22457	-	5.00007471B	PPB		27 Sep 2018 14:50:48	S	Std
7.5 - 1	57.5	32310	-	7.50007471B	PPB		27 Sep 2018 14:53:06	S	Std
10.0 - 1	510	42559	-	10.00007471B	PPB		27 Sep 2018 14:58:02	S	Std
ICV85 - 1	ICV85	17139	3.9633	-7471B	PPB		27 Sep 2018 15:01:02	U	SMPL
ICB85 - 1	ICB85	47	-0.0308	-7471B	PPB		27 Sep 2018 15:03:18	U	SMPL
CCV10 - 1	CCV10	20839	4.8279	-7471B	PPB		27 Sep 2018 15:05:34	U	SMPL
CCB10 - 1	CCB10	40	-0.0325	-7471B	PPB		27 Sep 2018 15:07:50	U	SMPL
CRA - 1	CRA	1068	0.2078	-7471B	PPB		27 Sep 2018 15:10:07	U	SMPL
HighStd - 1	HighStd	43039	10.0157	-7471B	PPB		27 Sep 2018 15:12:23	U	SMPL
ChkStd - 1	ChkStd	29535	6.8601	-7471B	PPB		27 Sep 2018 15:19:51	U	SMPL
PB113265BL - 1	PBS	-73	-0.0589	-7471B	PPB		27 Sep 2018 15:22:08	U	SMPL
PB113265BS - 1	LCSS	15108	3.4887	-7471B	PPB		27 Sep 2018 15:24:26	U	SMPL
J4773-11 - 1	CL-02-092518-A	2816	0.6163	-7471B	PPB		27 Sep 2018 15:26:43	U	SMPL
J4773-13 - 1	CL-02-092518-B	4485	1.0063	-7471B	PPB		27 Sep 2018 15:29:00	U	SMPL
J4773-15 - 1	CL-02-092518-C	1706	0.3569	-7471B	PPB		27 Sep 2018 15:31:18	U	SMPL
J4773-17 - 1	CL-02-092518-D	1097	0.2146	-7471B	PPB		27 Sep 2018 15:33:35	U	SMPL
J4773-19 - 1	CL-02-092518-E	1288	0.2592	-7471B	PPB		27 Sep 2018 15:35:51	U	SMPL
CCV11 - 1	CCV11	20497	4.7480	-7471B	PPB		27 Sep 2018 15:38:07	U	SMPL
CCB11 - 1	CCB11	-17	-0.0458	-7471B	PPB		27 Sep 2018 15:40:23	U	SMPL
J5074-01 - 1	CA-SOIL-1	706	0.1232	-7471B	PPB		27 Sep 2018 15:42:39	U	SMPL
J5074-11 - 1	CA-SOIL-2	1080	0.2106	-7471B	PPB		27 Sep 2018 15:44:56	U	SMPL
J5077-01 - 1	PARCEL-186-092418-1	4743	1.0666	-7471B	PPB		27 Sep 2018 15:47:12	U	SMPL
J5077-02 - 1	PARCEL-186-092418-2	4514	1.0130	-7471B	PPB		27 Sep 2018 15:49:29	U	SMPL
J5077-03 - 1	PARCEL-190-092418-1	2304	0.4966	-7471B	PPB		27 Sep 2018 15:51:45	U	SMPL
J5077-04 - 1	PARCEL-190-092418-2	1999	0.4253	-7471B	PPB		27 Sep 2018 15:54:02	U	SMPL
J5077-05 - 1	PARCEL-193-092418-1	4729	1.0633	-7471B	PPB		27 Sep 2018 15:56:19	U	SMPL
J5077-05DUP - 1	PARCEL-193-092418-1DUP	5116	1.1537	-7471B	PPB		27 Sep 2018 15:58:37	U	SMPL
J5077-05MS - 1	PARCEL-193-092418-1MS	20594	4.7707	-7471B	PPB		27 Sep 2018 16:00:54	U	SMPL
J5077-05MSD - 1	PARCEL-193-092418-1MSD	20885	4.8387	-7471B	PPB		27 Sep 2018 16:03:10	U	SMPL
CCV12 - 1	CCV12	20789	4.8163	-7471B	PPB		27 Sep 2018 16:05:27	U	SMPL
CCB12 - 1	CCB12	51	-0.0299	-7471B	PPB		27 Sep 2018 16:07:44	U	SMPL
J5077-06 - 1	PARCEL-193-092418-2	3195	0.7048	-7471B	PPB		27 Sep 2018 16:10:02	U	SMPL
J5081-03 - 1	TP-1-S	5637	1.2755	-7471B	PPB		27 Sep 2018 16:12:18	U	SMPL
J5081-07 - 1	TP-1-D	1228	0.2452	-7471B	PPB		27 Sep 2018 16:14:34	U	SMPL
J5081-11 - 1	TP-2-S	10675	2.4528	-7471B	PPB		27 Sep 2018 16:16:51	U	SMPL
J5081-15 - 1	TP-2-D	661	0.1127	-7471B	PPB		27 Sep 2018 16:19:07	U	SMPL
J5084-21 - 1	B-13(1-1.5)	2738	0.5980	-7471B	PPB		27 Sep 2018 16:21:24	U	SMPL
J5093-01 - 1	TOP-SOIL	2310	0.4980	-7471B	PPB		27 Sep 2018 16:23:41	U	SMPL
PB113267BL - 1	PBS	106	-0.0170	-7471B	PPB		27 Sep 2018 16:25:59	U	SMPL
PB113267BS - 1	LCSS	16076	3.7149	-7471B	PPB		27 Sep 2018 16:28:17	U	SMPL
J5066-01 - 1	EP1-Q1-HI1	2988	0.6564	-7471B	PPB		27 Sep 2018 16:30:33	U	SMPL
CCV13 - 1	CCV13	20979	4.8607	-7471B	PPB		27 Sep 2018 16:32:49	U	SMPL
CCB13 - 1	CCB13	-7	-0.0434	-7471B	PPB		27 Sep 2018 16:35:06	U	SMPL
J5066-02 - 1	EP1-Q1-G1	1741	0.3650	-7471B	PPB		27 Sep 2018 16:37:23	U	SMPL
J5066-03 - 1	EP1-Q1-G2	1252	0.2508	-7471B	PPB		27 Sep 2018 16:39:39	U	SMPL
J5066-04 - 1	EP1-Q1-G3	1033	0.1996	-7471B	PPB		27 Sep 2018 16:41:56	U	SMPL
J5066-04DUP - 1	EP1-Q1-G3DUP	1101	0.2155	-7471B	PPB		27 Sep 2018 16:44:12	U	SMPL
J5066-05 - 1	J5066-04MS	17394	4.0229	-7471B	PPB		27 Sep 2018 16:46:29	U	SMPL
J5066-06 - 1	J5066-04MSD	17468	4.0402	-7471B	PPB		27 Sep 2018 16:48:47	U	SMPL
J5066-07 - 1	EP1-Q1-F3	2119	0.4534	-7471B	PPB		27 Sep 2018 16:51:05	U	SMPL
J5066-08 - 1	EP1-Q1-F4	979	0.1870	-7471B	PPB		27 Sep 2018 16:53:23	U	SMPL
J5066-09 - 1	EP1-Q1-H4	1324	0.2676	-7471B	PPB		27 Sep 2018 16:55:40	U	SMPL
J5066-10 - 1	EP1-Q1-I4	1663	0.3468	-7471B	PPB		27 Sep 2018 16:57:57	U	SMPL
CCV14 - 1	CCV14	22614	5.2427	-7471B	PPB		27 Sep 2018 17:00:13	U	SMPL
CCB14 - 1	CCB14	55	-0.0289	-7471B	PPB		27 Sep 2018 17:02:30	U	SMPL
J5066-11 - 1	EP1-Q1-H3	2215	0.4758	-7471B	PPB		27 Sep 2018 17:04:47	U	SMPL
J5066-12 - 1	EP1-Q1-I3	2043	0.4356	-7471B	PPB		27 Sep 2018 17:07:04	U	SMPL
J5066-13 - 1	EP1-Q1-H2	1475	0.3029	-7471B	PPB		27 Sep 2018 17:09:20	U	SMPL
J5066-14 - 1	EP1-Q1-I2	1361	0.2762	-7471B	PPB		27 Sep 2018 17:11:36	U	SMPL
J5066-15 - 1	EP1-Q1-FD01	3933	0.8773	-7471B	PPB		27 Sep 2018 17:13:53	U	SMPL
PB113268BL - 1	PBS	113	-0.0154	-7471B	PPB		27 Sep 2018 17:16:10	U	SMPL
PB113268BS - 1	LCSS	14951	3.4520	-7471B	PPB		27 Sep 2018 17:18:27	U	SMPL
J5096-01 - 1	MT-T-5-2	2584	0.5620	-7471B	PPB		27 Sep 2018 17:20:44	U	SMPL
J5097-01 - 1	TP-1	792	0.1433	-7471B	PPB		27 Sep 2018 17:23:03	U	SMPL
J5097-01DUP - 1	TP-1DUP	752	0.1339	-7471B	PPB		27 Sep 2018 17:25:19	U	SMPL
CCV15 - 1	CCV15	22537	5.2247	-7471B	PPB		27 Sep 2018 17:27:36	U	SMPL
CCB15 - 1	CCB15	69	-0.0257	-7471B	PPB		27 Sep 2018 17:29:53	U	SMPL
J5097-01MS - 1	TP-1MS	17055	3.9437	-7471B	PPB		27 Sep 2018 17:32:12	U	SMPL
J5097-01MSD - 1	TP-1MSD	17429	4.0311	-7471B	PPB		27 Sep 2018 17:34:29	U	SMPL
J5097-03 - 1	TP-2	813	0.1482	-7471B	PPB		27 Sep 2018 17:36:46	U	SMPL
J5097-05 - 1	TP-3	2105	0.4501	-7471B	PPB		27 Sep 2018 17:39:04	U	SMPL
J5097-07 - 1	TP-4	2528	0.5490	-7471B	PPB		27 Sep 2018 17:41:20	U	SMPL
J5099-01 - 1	7211975	1628	0.3386	-7471B	PPB		27 Sep 2018 17:43:37	U	SMPL
J5100-01 - 1	HS-STONE-RO-227	36974	8.5984	-7471B	PPB		27 Sep 2018 17:45:54	U	SMPL
J5102-01 - 1	C-1	1174	0.2325	-7471B	PPB		27 Sep 2018 17:48:12	U	SMPL
J5105-01 - 1	WJ-18-09-26-TSHG-09	2342	0.5055	-7471B	PPB		27 Sep 2018 17:50:32	U	SMPL
J5105-02 - 1	WJ-18-09-26-TSHG-10	2398	0.5186	-7471B	PPB		27 Sep 2018 17:52:49	U	SMPL
CCV16 - 1	CCV16	21024	4.8712	-7471B	PPB		27 Sep 2018 17:55:06	U	SMPL

LB98188 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Sep 2018 14:36:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
CCB16 - 1	CCB16	71	-0.0252		-7471B	PPB	27 Sep 2018 17:57:23	U	SMPL
J5105-03 - 1	WJ-18-09-26-TSHG-11	2324	0.5013		-7471B	PPB	27 Sep 2018 17:59:41	U	SMPL
J5077-05LX5 - 1		1075	0.2094		-7471B	PPB	27 Sep 2018 18:01:58	U	SMPL
J5077-05A - 1		22221	5.1509		-7471B	PPB	27 Sep 2018 18:04:14	U	SMPL
J5066-04LX5 - 1		211	0.0075		-7471B	PPB	27 Sep 2018 18:06:31	U	SMPL
J5066-04A - 1		18869	4.3676		-7471B	PPB	27 Sep 2018 18:08:48	U	SMPL
J5097-01LX5 - 1		240	0.0143		-7471B	PPB	27 Sep 2018 18:11:05	U	SMPL
J5097-01A - 1		18456	4.2711		-7471B	PPB	27 Sep 2018 18:13:22	U	SMPL
CCV17 - 1	CCV17	21420	4.9637		-7471B	PPB	27 Sep 2018 18:15:40	U	SMPL
CCB17 - 1	CCB17	109	-0.0163		-7471B	PPB	27 Sep 2018 18:17:59	U	SMPL