


ISM02.4

LB98215
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45295



A= 0.0000e+000

B= 2.2160e-004

C= 8.8048e-003

Rho= 0.9999465

Accept=Accepted

slope
 $y = 1.01 \times 10^4 x$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.042	0.042	148	0.000	148					1
0.2	0.200	0.221	0.021	956	0.0 %	956					1
2.5	2.500	2.456	-0.044	11043	0.0 %	11043					2
5.0	5.000	4.953	-0.047	22310	0.0 %	22310					1
7.5	7.500	7.483	-0.017	33730	0.0 %	33730					0
10.0	10.000	10.046	0.046	45295	0.0 %	45295					0

NB
9/28/18

LB98215
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	148	-		0 ISM02.4	PPB	9/28/2018 9:00	Std
	0.2 S0.2	956	-		0.2 ISM02.4	PPB	9/28/2018 9:02	Std
	2.5 S2.5	11043	-		2.5 ISM02.4	PPB	9/28/2018 9:05	Std
	5 S5	22310	-		5 ISM02.4	PPB	9/28/2018 9:07	Std
	7.5 S7.5	33730	-		7.5 ISM02.4	PPB	9/28/2018 9:09	Std
	10 S10.0	45295	-		10 ISM02.4	PPB	9/28/2018 9:12	Std
ICV	ICV	18135	4.0275 -		ISM02.4	PPB	9/28/2018 9:14	SMPL
ICB	ICB	32	0.0159 -		ISM02.4	PPB	9/28/2018 9:17	SMPL
CCV021	CCV021	22813	5.0641 -		ISM02.4	PPB	9/28/2018 9:19	SMPL
CCB021	CCB021	10	0.011 -		ISM02.4	PPB	9/28/2018 9:21	SMPL
PB113298BL	PBS004	31	0.0157 -		ISM02.4	PPB	9/28/2018 9:23	SMPL
J4867-01	MYAKD5	1761	0.399 -		ISM02.4	PPB	9/28/2018 9:26	SMPL
J4867-02	MYAKD6	2952	0.663 -		ISM02.4	PPB	9/28/2018 9:28	SMPL
J4867-03	MYAKD7	3646	0.8167 -		ISM02.4	PPB	9/28/2018 9:30	SMPL
J4867-04	MYAKD8	3373	0.7562 -		ISM02.4	PPB	9/28/2018 9:33	SMPL
J4867-05	MYAKD9	2867	0.6441 -		ISM02.4	PPB	9/28/2018 9:35	SMPL
J4867-06	MYAKD9D	3019	0.6778 -		ISM02.4	PPB	9/28/2018 9:37	SMPL
J4867-07	MYAKD9S	15651	3.477 -		ISM02.4	PPB	9/28/2018 9:39	SMPL
J4867-08	MYAKF0	8602	1.915 -		ISM02.4	PPB	9/28/2018 9:42	SMPL
J4867-09	MYAKF1	11018	2.4504 -		ISM02.4	PPB	9/28/2018 9:44	SMPL
J4867-10	MYAKF2	4238	0.9479 -		ISM02.4	PPB	9/28/2018 9:46	SMPL
J4867-11	MYAKF3	3539	0.793 -		ISM02.4	PPB	9/28/2018 9:49	SMPL
J4867-12	MYAKF4	4192	0.9377 -		ISM02.4	PPB	9/28/2018 9:51	SMPL
J4867-13	MYAKF5	4731	1.0572 -		ISM02.4	PPB	9/28/2018 9:53	SMPL
J4867-14	MYAKF6	4619	1.0324 -		ISM02.4	PPB	9/28/2018 9:55	SMPL
J4867-15	MYAKF7	3377	0.7571 -		ISM02.4	PPB	9/28/2018 9:58	SMPL
J4867-16	MYAKF8	2959	0.6645 -		ISM02.4	PPB	9/28/2018 10:00	SMPL
J4867-17	MYAKF9	3559	0.7975 -		ISM02.4	PPB	9/28/2018 10:02	SMPL
J4867-18	MYAKG0	3593	0.805 -		ISM02.4	PPB	9/28/2018 10:04	SMPL
J4867-19	MYAKG1	3116	0.6993 -		ISM02.4	PPB	9/28/2018 10:07	SMPL
CCV022	CCV022	23342	5.1813 -		ISM02.4	PPB	9/28/2018 10:09	SMPL
CCB022	CCB022	-12	0.0061 -		ISM02.4	PPB	9/28/2018 10:11	SMPL
J4867-20	MYAKG2	3748	0.8393 -		ISM02.4	PPB	9/28/2018 10:14	SMPL
J4867-21	MYAKG3	4759	1.0634 -		ISM02.4	PPB	9/28/2018 10:16	SMPL
J4867-22	MYAKG4	6039	1.347 -		ISM02.4	PPB	9/28/2018 10:18	SMPL
PB113299BL	PBS004	46	0.019 -		ISM02.4	PPB	9/28/2018 10:20	SMPL
J4971-01	MYAKB3	3402	0.7627 -		ISM02.4	PPB	9/28/2018 10:23	SMPL
J4971-02	MYAKB4	3493	0.7828 -		ISM02.4	PPB	9/28/2018 10:25	SMPL
J4971-03	MYAKB5	7208	1.6061 -		ISM02.4	PPB	9/28/2018 10:27	SMPL
J4971-04	MYAKB6	7824	1.7426 -		ISM02.4	PPB	9/28/2018 10:30	SMPL
J4971-05	MYAKB7	8331	1.8549 -		ISM02.4	PPB	9/28/2018 10:32	SMPL
J4971-06	MYAKB8	4264	0.9537 -		ISM02.4	PPB	9/28/2018 10:34	SMPL
J4971-07	MYAKB9	5387	1.2025 -		ISM02.4	PPB	9/28/2018 10:36	SMPL
J4971-08	MYAKC0	5218	1.1651 -		ISM02.4	PPB	9/28/2018 10:39	SMPL
J4971-09	MYAKC1	6405	1.4281 -		ISM02.4	PPB	9/28/2018 10:41	SMPL
J4971-10	MYAKC2	7909	1.7614 -		ISM02.4	PPB	9/28/2018 10:43	SMPL

INSTRUMENT ID : CV1

J4971-11	MYAKC3	10817	2.4058 -	ISM02.4	PPB	9/28/2018 10:46	SMPL
J4971-12	MYAKC4	13336	2.964 -	ISM02.4	PPB	9/28/2018 10:48	SMPL
J4971-13	MYAKC5	12660	2.8142 -	ISM02.4	PPB	9/28/2018 10:50	SMPL
J4971-14	MYAKC6	17785	3.9499 -	ISM02.4	PPB	9/28/2018 10:52	SMPL
J4971-15	MYAKC9	1786	0.4046 -	ISM02.4	PPB	9/28/2018 10:55	SMPL
J4971-16	MYAKD0	3270	0.7334 -	ISM02.4	PPB	9/28/2018 10:57	SMPL
CCV023	CCV023	23379	5.1895 -	ISM02.4	PPB	9/28/2018 10:59	SMPL
CCB023	CCB023	-24	0.0035 -	ISM02.4	PPB	9/28/2018 11:02	SMPL
J4971-17	MYAKD1	3724	0.834 -	ISM02.4	PPB	9/28/2018 11:04	SMPL
J4971-18	MYAKD2	3991	0.8932 -	ISM02.4	PPB	9/28/2018 11:06	SMPL
J4971-19	MYAKD3	4413	0.9867 -	ISM02.4	PPB	9/28/2018 11:08	SMPL
J4971-20	MYAKD4	4008	0.897 -	ISM02.4	PPB	9/28/2018 11:11	SMPL
J4971-21	MYAKD4D	4143	0.9269 -	ISM02.4	PPB	9/28/2018 11:13	SMPL
J4971-22	MYAKD4S	16374	3.6372 -	ISM02.4	PPB	9/28/2018 11:15	SMPL
PB113300BL	PBS004	135	0.0387 -	ISM02.4	PPB	9/28/2018 11:17	SMPL
J4972-01	MYAK56	669	0.1571 -	ISM02.4	PPB	9/28/2018 11:20	SMPL
J4972-02	MYAK57	842	0.1954 -	ISM02.4	PPB	9/28/2018 11:22	SMPL
J4972-03	MYAK90	7232	1.6114 -	ISM02.4	PPB	9/28/2018 11:24	SMPL
J4972-04	MYAK91	24644	5.4698 -	ISM02.4	PPB	9/28/2018 11:27	SMPL
J4972-05	MYAK92	5299	1.183 -	ISM02.4	PPB	9/28/2018 11:29	SMPL
J4972-06	MYAK93	1292	0.2951 -	ISM02.4	PPB	9/28/2018 11:31	SMPL
J4972-07	MYAK94	1655	0.3755 -	ISM02.4	PPB	9/28/2018 11:33	SMPL
J4972-08	MYAK99	3554	0.7964 -	ISM02.4	PPB	9/28/2018 11:36	SMPL
J4972-09	MYAKA0	6694	1.4922 -	ISM02.4	PPB	9/28/2018 11:38	SMPL
J4972-10	MYAKA1	6660	1.4846 -	ISM02.4	PPB	9/28/2018 11:40	SMPL
J4972-11	MYAKA2	5937	1.3244 -	ISM02.4	PPB	9/28/2018 11:43	SMPL
J4972-12	MYAKA3	7421	1.6533 -	ISM02.4	PPB	9/28/2018 11:45	SMPL
J4972-13	MYAKA4	4776	1.0671 -	ISM02.4	PPB	9/28/2018 11:47	SMPL
CCV024	CCV024	23287	5.1691 -	ISM02.4	PPB	9/28/2018 11:49	SMPL
CCB024	CCB024	-27	0.0028 -	ISM02.4	PPB	9/28/2018 11:52	SMPL
J4972-14	MYAKA5	5101	1.1392 -	ISM02.4	PPB	9/28/2018 11:54	SMPL
J4972-15	MYAKA6	4926	1.1004 -	ISM02.4	PPB	9/28/2018 11:56	SMPL
J4972-16	MYAKA8	2677	0.602 -	ISM02.4	PPB	9/28/2018 11:59	SMPL
J4972-17	MYAKA9	4323	0.9668 -	ISM02.4	PPB	9/28/2018 12:01	SMPL
J4972-18	MYAKB0	3901	0.8733 -	ISM02.4	PPB	9/28/2018 12:03	SMPL
J4972-19	MYAKB1	5953	1.328 -	ISM02.4	PPB	9/28/2018 12:05	SMPL
J4972-20	MYAKB2	4485	1.0027 -	ISM02.4	PPB	9/28/2018 12:08	SMPL
J4972-21	MYAKB2D	4347	0.9721 -	ISM02.4	PPB	9/28/2018 12:10	SMPL
J4972-22	MYAKB2S	16521	3.6698 -	ISM02.4	PPB	9/28/2018 12:12	SMPL
PB113301BL	PBS004	64	0.023 -	ISM02.4	PPB	9/28/2018 12:15	SMPL
J4976-01	MYAKG5	7724	1.7204 -	ISM02.4	PPB	9/28/2018 12:17	SMPL
J4976-02	MYAKG6	10501	2.3358 -	ISM02.4	PPB	9/28/2018 12:19	SMPL
J4976-03	MYAKH2	5657	1.2624 -	ISM02.4	PPB	9/28/2018 12:22	SMPL
J4976-04	MYAKH3	8985	1.9998 -	ISM02.4	PPB	9/28/2018 12:24	SMPL
J4976-05	MYAKH4	3273	0.7341 -	ISM02.4	PPB	9/28/2018 12:26	SMPL
J4976-06	MYAKH5	3212	0.7206 -	ISM02.4	PPB	9/28/2018 12:28	SMPL
J4976-07	MYAKH6	2769	0.6224 -	ISM02.4	PPB	9/28/2018 12:31	SMPL

INSTRUMENT ID : CV1

J4976-08	MYAKH7	4533	1.0133 -	ISM02.4	PPB	9/28/2018 12:33	SMPL
J4976-09	MYAKH8	5225	1.1666 -	ISM02.4	PPB	9/28/2018 12:35	SMPL
J4976-10	MYAKH9	2390	0.5384 -	ISM02.4	PPB	9/28/2018 12:38	SMPL
CCV025	CCV025	23492	5.2145 -	ISM02.4	PPB	9/28/2018 12:40	SMPL
CCB025	CCB025	79	0.0263 -	ISM02.4	PPB	9/28/2018 12:42	SMPL
J4976-11	MYAKJ0	3209	0.7199 -	ISM02.4	PPB	9/28/2018 12:44	SMPL
J4976-12	MYAKJ1	3406	0.7636 -	ISM02.4	PPB	9/28/2018 12:47	SMPL
J4976-13	MYAKJ2	2800	0.6293 -	ISM02.4	PPB	9/28/2018 12:49	SMPL
J4976-14	MYAKJ3	2436	0.5486 -	ISM02.4	PPB	9/28/2018 12:51	SMPL
J4976-15	MYAKJ4	3866	0.8655 -	ISM02.4	PPB	9/28/2018 12:54	SMPL
J4976-16	MYAKJ5	4513	1.0089 -	ISM02.4	PPB	9/28/2018 12:56	SMPL
J4976-17	MYAKJ6	6294	1.4035 -	ISM02.4	PPB	9/28/2018 12:58	SMPL
J4976-18	MYAKJ6S	18420	4.0906 -	ISM02.4	PPB	9/28/2018 13:00	SMPL
J4976-19	MYAKJ6D	6572	1.4651 -	ISM02.4	PPB	9/28/2018 13:03	SMPL
J4976-20	MYAKJ7	9124	2.0306 -	ISM02.4	PPB	9/28/2018 13:05	SMPL
J4976-21	MYAKJ8	18252	4.0534 -	ISM02.4	PPB	9/28/2018 13:07	SMPL
J4976-22	MYAKP0	950	0.2193 -	ISM02.4	PPB	9/28/2018 13:10	SMPL
PB113302BL	PBS004	21	0.0135 -	ISM02.4	PPB	9/28/2018 13:12	SMPL
J4977-01	MYAKP3	5044	1.1265 -	ISM02.4	PPB	9/28/2018 13:14	SMPL
J4977-02	MYAKP4	9616	2.1397 -	ISM02.4	PPB	9/28/2018 13:17	SMPL
J4977-03	MYAKY5	6725	1.499 -	ISM02.4	PPB	9/28/2018 13:19	SMPL
J4977-04	MYAKY6	3640	0.8154 -	ISM02.4	PPB	9/28/2018 13:21	SMPL
J4977-05	MYAKY7	4870	1.088 -	ISM02.4	PPB	9/28/2018 13:23	SMPL
CCV026	CCV026	23469	5.2094 -	ISM02.4	PPB	9/28/2018 13:26	SMPL
CCB026	CCB026	81	0.0268 -	ISM02.4	PPB	9/28/2018 13:28	SMPL
J4977-06	MYAKY8	4774	1.0667 -	ISM02.4	PPB	9/28/2018 13:30	SMPL
J4977-07	MYAKY8D	4610	1.0304 -	ISM02.4	PPB	9/28/2018 13:33	SMPL
J4977-08	MYAKY8S	16551	3.6764 -	ISM02.4	PPB	9/28/2018 13:35	SMPL
CCV027	CCV027	23300	5.172 -	ISM02.4	PPB	9/28/2018 13:37	SMPL
CCB027	CCB027	48	0.0194 -	ISM02.4	PPB	9/28/2018 13:39	SMPL