

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: AMANDEEP

ANALYST: Aparna

Date: 09/28/2018

TEMP1 IN: 103 09/28/2018 07:00 TEMP OUT: 104 09/28/2018 08:10

TEMP2 IN: 104 09/28/2018 09:00 TEMP OUT: 104 09/28/2018 10:05

TEMP3 IN: 105 09/28/2018 11:00 TEMP OUT: 105 09/28/2018 12:35

TEMP4 IN: 105 09/28/2018 13:45 TEMP OUT: 105 09/28/2018 15:20

Run Number: LB98217

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB98217BL	LB98217BL	1.3912	1.3912	100	1.3912	1.3912	1.3912	0.0000	0
2	LB98217BS	LB98217BS	1.3952	1.3952	100	1.4453	1.4453	1.4453	0.0501	501
3	J5142-01	SPDES-1	1.3912	1.3912	900	1.3938	1.3938	1.3938	0.0026	2.9

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS092818-3

WorkList ID : 117326

Date : 9/28/2018 9:52:51 AM

2898217

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/28/2018	Water	J5142-01 f	TSS	Cool 4 deg C	SKAN06	M43	SPDES-1	09/27/2018	SM2540 D

Date/Time 09/28/18 10:00
 Received by: AS
 Relinquished by: AS

Date/Time 09/28/18 12:00
 Received by: AS
 Relinquished by: AS