



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: Aparna

Run Number: LB98231

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP68201
Turbidity std, 40NTU	WP68196
turbidity std 5 NTU	WP68199
turbidity std 20 NTU	WP68197
Turbidity std 1 NTU	WP68200
turbidity std 0 NTU	WP68195
turbidity std 10 NTU	WP68198
10 NTU Standard 500 ml	W2297

Intercept: 0.2888

Slope: 0.9146

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.275	09/28/2018	13:21
2	CAL2	1	1	0.991	09/28/2018	13:25
3	CAL3	5	1	4.860	09/28/2018	13:30
4	CAL4	10	1	9.720	09/28/2018	13:35
5	CAL5	20	1	18.600	09/28/2018	13:40
6	CAL6	40	1	36.800	09/28/2018	13:45



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: Aparna

Run Number: LB98231

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.360	9.918	09/28/2018	13:53
2	ICB		1	0.276	-0.014	09/28/2018	13:58
3	CCV1	10	1	9.500	10.071	09/28/2018	14:05
4	CCB1		1	0.297	0.009	09/28/2018	14:00
5	LB98231BL		1	0.285	-0.004	09/28/2018	14:11
6	J5162-01		1	9.230	9.776	09/28/2018	14:15
7	J5162-01DUP		1	9.220	9.765	09/28/2018	14:17
8	CCV2	10	1	9.480	10.049	09/28/2018	14:22
9	CCB2		1	0.286	-0.003	09/28/2018	14:27

WORKLIST(Hardcopy Internal Chain)

LB 98231

WorkList Name : turbidity092818

WorkList ID : 117337

Date : 9/28/2018 11:30:18 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/29/2018	Water	J5162-01	Turbidity	Cool 4 deg C	LIRO01	M41	UDS-ER22-01	09/27/2018	SM2130 B

9-26-18 13:00
 LB 9823
 AC
 Received by: AC
 Relinquished by: AC

9-26-18 15:00
 Received by: AC
 Relinquished by: AC