



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98234

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	09/28/2018	11:30
2	CAL2	1	Water			22.4	7.01	09/28/2018	11:32
3	CAL3	1	Water			22.3	10.04	09/28/2018	11:34
4	ICV	1	Water			22.4	7.00	09/28/2018	11:36
5	CCV1	1	Water			22.2	2.00	09/28/2018	11:38
6	J5145-01	1	Solid	20.02	20	23.2	5.10	09/28/2018	11:40
7	J5145-01DUP	1	Solid	20.04	20	23.3	5.11	09/28/2018	11:42
8	CCV2	1	Water			22.2	12.00	09/28/2018	11:44

WORKLIST(Hardcopy Internal Chain)

LB98 234

WorkList Name : pH092818

WorkList ID : 117335

Date : 9/28/2018 11:19:30 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/27/2018	Solid	J5145-01	pH	Cool 4 deg C	TULL02	M33	FK-PS-01-042	09/27/2018	9045D

Date/Time 09/28/18 11:20
 Received by: SP
 Relinquished by: CP

Date/Time 09/28/18 11:50
 Received by: CP
 Relinquished by: SP