

LA 98240


7471B

INSTRUMENT ID: CVI

Linear

μ Abs.:

41301



A= 0.0000e+000

B= 2.4147e-004

C= 1.1198e-002

Rho= 0.9998871

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.045	0.045	139	0.000	139				
0.2	0.200	0.199	-0.001	777	0.0 %	777				
2.5	2.500	2.395	-0.105	9871	0.0 %	9871				
5.0	5.000	5.071	0.071	20955	0.0 %	20955				
7.5	7.500	7.507	0.007	31041	0.0 %	31041				
10.0	10.000	9.984	-0.016	41301	0.0 %	41301				

LB98240 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Sep 2018 18:22:24

Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	SD	139	-	0.00007471B	PPB	01 Oct 2018 09:02:16	Std	S
0.2 - 1	SD	777	-	0.20007471B	PPB	01 Oct 2018 09:04:34	Std	S
2.5 - 1	SD	9871	-	2.50007471B	PPB	01 Oct 2018 09:06:52	Std	S
5.0 - 1	SD	20955	-	5.00007471B	PPB	01 Oct 2018 09:09:10	Std	S
7.5 - 1	SD	31041	-	7.50007471B	PPB	01 Oct 2018 09:11:27	Std	S
10.0 - 1	SD	41301	-	10.00007471B	PPB	01 Oct 2018 09:16:16	Std	S
ICV89 - 1	ICV89	16814	4.0712	-7471B	PPB	01 Oct 2018 09:24:10	SMPL	U
ICB89 - 1	ICB89	-76	-0.0072	-7471B	PPB	01 Oct 2018 09:26:26	SMPL	U
CCV33 - 1	CCV33	20957	5.0716	-7471B	PPB	01 Oct 2018 09:28:42	SMPL	U
CCB33 - 1	CCB33	-96	-0.0120	-7471B	PPB	01 Oct 2018 09:30:58	SMPL	U
CRA - 1	CRA	784	0.2005	-7471B	PPB	01 Oct 2018 09:33:15	SMPL	U
HighStd - 1	HighStd	41622	10.0615	-7471B	PPB	01 Oct 2018 09:35:31	SMPL	U
ChkStd - 1	ChkStd	29174	7.0557	-7471B	PPB	01 Oct 2018 09:37:47	SMPL	U
PB113321BL - 1	PBS	-182	-0.0327	-7471B	PPB	01 Oct 2018 09:40:05	SMPL	U
PB113321BS - 1	LCSS	16443	3.9816	-7471B	PPB	01 Oct 2018 09:42:25	SMPL	U
J4771-05 - 1	NB-08-092718	1155	0.2901	-7471B	PPB	01 Oct 2018 09:49:40	SMPL	U
J4771-05DUP - 1	NB-08-092718DUP	1036	0.2614	-7471B	PPB	01 Oct 2018 09:51:56	SMPL	U
J4771-05MS - 1	NB-08-092718MS	17183	4.1603	-7471B	PPB	01 Oct 2018 09:54:14	SMPL	U
J4771-05MSD - 1	NB-08-092718MSD	17228	4.1712	-7471B	PPB	01 Oct 2018 09:56:31	SMPL	U
J4776-05 - 1	EO-01-092718-A	2699	0.6629	-7471B	PPB	01 Oct 2018 09:58:47	SMPL	U
CCV34 - 1	CCV34	20767	5.0257	-7471B	PPB	01 Oct 2018 10:01:04	SMPL	U
CCB34 - 1	CCB34	440	0.1174	-7471B	PPB	01 Oct 2018 10:03:20	SMPL	U
J4776-07 - 1	EO-01-092718-B	2227	0.5489	-7471B	PPB	01 Oct 2018 10:05:37	SMPL	U
J4776-09 - 1	EO-01-092718-C	1928	0.4767	-7471B	PPB	01 Oct 2018 10:07:52	SMPL	U
J4777-33 - 1	PL-01-092718-A	2565	0.6306	-7471B	PPB	01 Oct 2018 10:10:08	SMPL	U
J4777-35 - 1	PL-01-092718-B	1252	0.3135	-7471B	PPB	01 Oct 2018 10:12:25	SMPL	U
J4777-37 - 1	PL-01-092718-C	2580	0.6342	-7471B	PPB	01 Oct 2018 10:14:41	SMPL	U
J4777-39 - 1	PL-01-092718-D	1653	0.4103	-7471B	PPB	01 Oct 2018 10:16:58	SMPL	U
J5118-01 - 1	RA-092518-USTEXC-NW-L-063	44	0.0218	-7471B	PPB	01 Oct 2018 10:19:16	SMPL	U
J5118-02 - 1	RA-092518-USTEXC-NW-R-064	2853	0.7001	-7471B	PPB	01 Oct 2018 10:21:34	SMPL	U
J4771-05LX5 - 1		280	0.0788	-7471B	PPB	01 Oct 2018 10:26:48	SMPL	U
J4771-05A - 1		19687	4.7650	-7471B	PPB	01 Oct 2018 10:29:03	SMPL	U
CCV35 - 1	CCV35	20859	5.0480	-7471B	PPB	01 Oct 2018 10:31:20	SMPL	U
CCB35 - 1	CCB35	130	0.0426	-7471B	PPB	01 Oct 2018 10:36:10	SMPL	U