



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 10/01/2018

TEMP1 IN: 103 10/01/2018 07:00 TEMP OUT: 104 10/01/2018 08:05

TEMP2 IN: 104 10/01/2018 09:00 TEMP OUT: 104 10/01/2018 10:00

TEMP3 IN: 105 10/01/2018 11:00 TEMP OUT: 105 10/01/2018 12:35

TEMP4 IN: 105 10/01/2018 13:15 TEMP OUT: 105 10/01/2018 15:20

Run Number: LB98245

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB98245BL	LB98245BL	1.4102	1.4102	100	1.4102	1.4102	1.4102	0.0000	0
2	LB98245BS	LB98245BS	1.4000	1.4000	100	1.4525	1.4525	1.4525	0.0525	525
3	J5162-01	UDS-ER22-01	1.4016	1.4016	600	1.4046	1.4046	1.4046	0.0030	5

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS100118

WorkList ID : 117365

Date : 10/1/2018 7:49:21 AM

2098245

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2018	Water	J5162-01	TSS	Cool 4 deg C	LIR001	M41	UDS-ER22-01	09/27/2018	SM2540 D

FP-9823888

Date/Time 10/1/18, 08:00

Received by: AS

Relinquished by: [Signature]

Date/Time 10/1/18, 12:30

Received by: AS

Relinquished by: AS