



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98247

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	09/28/2018	17:00
2	CAL2	1	Water			22.0	7.01	09/28/2018	17:02
3	CAL3	1	Water			21.9	10.04	09/28/2018	17:04
4	ICV	1	Water			22.2	7.00	09/28/2018	17:06
5	CCV1	1	Water			21.9	2.00	09/28/2018	17:08
6	J4780-04	1	Solid	20.02	20	22.8	5.78	09/28/2018	17:10
7	J4780-04DUP	1	Solid	20.04	20	22.9	5.79	09/28/2018	17:12
8	J5154-03	1	Solid	20.03	20	23.2	8.55	09/28/2018	17:14
9	J5170-06	1	Solid	20.03	20	22.9	8.26	09/28/2018	17:16
10	J5170-12	1	Solid	20.02	20	22.9	8.22	09/28/2018	17:18
11	J5170-18	1	Solid	20.04	20	22.6	8.17	09/28/2018	17:20
12	J5170-24	1	Solid	20.02	20	22.6	8.02	09/28/2018	17:22
13	J5170-30	1	Solid	20.02	20	22.8	8.15	09/28/2018	17:24
14	J5171-04	1	Solid	20.03	20	23.1	8.80	09/28/2018	17:26
15	CCV2	1	Water			21.9	12.00	09/28/2018	17:28

WORKLIST(Hardcopy Internal Chain)

LB 98247

WorkList Name : CORROSIONITY092818

WorkList ID : 117358

Date : 9/28/2018 4:20:32 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/28/2018	Solid	J4780-04	Corrosivity	Cool 4 deg C	PSEG05		TR-04-092818	09/28/2018	9045D
09/28/2018	Solid	J5154-03	Corrosivity	Cool 4 deg C	PSEG01	J21	OILY-STONES	09/28/2018	9045D
09/28/2018	Solid	J5170-06	Corrosivity	Cool 4 deg C	PSEG01	J31	HC-SOIL-1	09/28/2018	9045D
09/28/2018	Solid	J5170-12	Corrosivity	Cool 4 deg C	PSEG01	J31	HC-SOIL-2	09/28/2018	9045D
09/28/2018	Solid	J5170-18	Corrosivity	Cool 4 deg C	PSEG01	J31	HC-SOIL-3	09/28/2018	9045D
09/28/2018	Solid	J5170-24	Corrosivity	Cool 4 deg C	PSEG01	J31	HC-SOIL-4	09/28/2018	9045D
09/28/2018	Solid	J5170-30	Corrosivity	Cool 4 deg C	PSEG01	J31	HC-SOIL-5	09/28/2018	9045D
09/28/2018	Solid	J5171-04	Corrosivity	Cool 4 deg C	PSEG01	J21	KR-RO-VNJ-221	09/28/2018	9045D

Date/Time 09/28/18 4:40

Received by: CP

Relinquished by: SP

Date/Time 09/28/18 5:35

Received by: CP

Relinquished by: SP