


7470A

LB 98256
INSTRUMENT ID: CY1

Linear

 μ Abs.:

36707



A= 0.0000e+000

B= 2.7088e-004

C= -1.3412e-002

Rho= 0.9999112

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	36	0.000	36				
0.2	0.200	0.161	-0.039	644	0.0 %	644				
2.5	2.500	2.491	-0.009	9246	0.0 %	9246				
5.0	5.000	5.078	0.078	18796	0.0 %	18796				
7.5	7.500	7.543	0.043	27897	0.0 %	27897				
10.0	10.000	9.930	-0.070	36707	0.0 %	36707				

LB98256 INSTRUMENT ID:CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Oct 2018 12:28:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50.0	36	-	0.0000	7470A	PPB	01 Oct 2018 12:29:45	S	Std
0.2 - 1	50.2	644	-	0.2000	7470A	PPB	01 Oct 2018 12:31:53	S	Std
2.5 - 1	52.5	9246	-	2.5000	7470A	PPB	01 Oct 2018 12:34:02	S	Std
5.0 - 1	55.0	18796	-	5.0000	7470A	PPB	01 Oct 2018 12:36:10	S	Std
7.5 - 1	57.5	27897	-	7.5000	7470A	PPB	01 Oct 2018 12:38:19	S	Std
10.0 - 1	510.0	36707	-	10.0000	7470A	PPB	01 Oct 2018 12:40:27	S	Std
ICV92 - 1	ICV92	14784	3.9913	-	7470A	PPB	01 Oct 2018 12:43:52	U	SMPL
ICB92 - 1	ICB92	-214	-0.0714	-	7470A	PPB	01 Oct 2018 12:45:59	U	SMPL
CCV36 - 1	CCV36	18662	5.0418	-	7470A	PPB	01 Oct 2018 12:48:05	U	SMPL
CCB36 - 1	CCB36	-161	-0.0570	-	7470A	PPB	01 Oct 2018 12:50:12	U	SMPL
CRA - 1	CRA	734	0.1854	-	7470A	PPB	01 Oct 2018 12:52:19	U	SMPL
HighStd - 1	HighStd	37029	10.0171	-	7470A	PPB	01 Oct 2018 12:54:26	U	SMPL
ChkStd - 1	ChkStd	26911	7.2763	-	7470A	PPB	01 Oct 2018 12:56:33	U	SMPL
PB113345BL - 1	PBW	-202	-0.0681	-	7470A	PPB	01 Oct 2018 12:58:42	U	SMPL
PB113345TB - 1	PB113305TB	-13	-0.0169	-	7470A	PPB	01 Oct 2018 13:02:58	U	SMPL
J4771-06 - 1	NB-08-092718	85	0.0096	-	7470A	PPB	01 Oct 2018 13:05:07	U	SMPL
J4771-06DUP - 1	NB-08-092718DUP	101	0.0139	-	7470A	PPB	01 Oct 2018 13:07:15	U	SMPL
CCV37 - 1	CCV37	18647	5.0378	-	7470A	PPB	01 Oct 2018 13:13:37	U	SMPL
CCB37 - 1	CCB37	-136	-0.0503	-	7470A	PPB	01 Oct 2018 13:15:44	U	SMPL
J4776-06 - 1	EO-01-092718-A	-2	-0.0140	-	7470A	PPB	01 Oct 2018 13:17:51	U	SMPL
J4776-08 - 1	EO-01-092718-B	44	-0.0015	-	7470A	PPB	01 Oct 2018 13:19:58	U	SMPL
J4776-10 - 1	EO-01-092718-C	92	0.0115	-	7470A	PPB	01 Oct 2018 13:22:04	U	SMPL
J4777-34 - 1	PL-01-092718-A	100	0.0137	-	7470A	PPB	01 Oct 2018 13:24:11	U	SMPL
J4777-36 - 1	PL-01-092718-B	24	-0.0069	-	7470A	PPB	01 Oct 2018 13:26:19	U	SMPL
J4777-38 - 1	PL-01-092718-C	66	0.0045	-	7470A	PPB	01 Oct 2018 13:28:26	U	SMPL
J4777-40 - 1	PL-01-092718-D	12	-0.0102	-	7470A	PPB	01 Oct 2018 13:30:35	U	SMPL
J5129-01 - 1	DS-DRUM-OILY-WIPE	46	-0.0010	-	7470A	PPB	01 Oct 2018 13:32:43	U	SMPL
PB113346BL - 1	PBW	10	-0.0107	-	7470A	PPB	01 Oct 2018 13:34:52	U	SMPL
PB113346BS - 1	LCSW	14533	3.9233	-	7470A	PPB	01 Oct 2018 13:39:23	U	SMPL
CCV38 - 1	CCV38	18492	4.9958	-	7470A	PPB	01 Oct 2018 13:41:30	U	SMPL
CCB38 - 1	CCB38	-128	-0.0481	-	7470A	PPB	01 Oct 2018 13:43:37	U	SMPL
J5148-01 - 1	WS-1	6316	1.6975	-	7470A	PPB	01 Oct 2018 13:45:46	U	SMPL
J5149-01 - 1	WS-2	1490	0.3902	-	7470A	PPB	01 Oct 2018 13:47:53	U	SMPL
J5149-01DUP - 1	WS-2DUP	1594	0.4184	-	7470A	PPB	01 Oct 2018 13:50:02	U	SMPL
J5149-01MS - 1	WS-2MS	16523	4.4624	-	7470A	PPB	01 Oct 2018 13:52:09	U	SMPL
J5149-01MSD - 1	WS-2MSD	16063	4.3378	-	7470A	PPB	01 Oct 2018 13:54:17	U	SMPL
J5149-02 - 1	WS-3	-56	-0.0286	-	7470A	PPB	01 Oct 2018 13:56:25	U	SMPL
PB113345BS - 1	LCSW	14758	3.9843	-	7470A	PPB	01 Oct 2018 13:58:34	U	SMPL
J4771-06MS - 1	NB-08-092718MS	14665	3.9591	-	7470A	PPB	01 Oct 2018 14:00:42	U	SMPL
J4771-06MSD - 1	NB-08-092718MSD	14299	3.8600	-	7470A	PPB	01 Oct 2018 14:02:52	U	SMPL
CCV39 - 1	CCV39	18364	4.9611	-	7470A	PPB	01 Oct 2018 14:04:59	U	SMPL
CCB39 - 1	CCB39	-142	-0.0519	-	7470A	PPB	01 Oct 2018 14:07:06	U	SMPL
J4771-06LX5 - 1		27	-0.0061	-	7470A	PPB	01 Oct 2018 14:09:15	U	SMPL
J4771-06A - 1		15346	4.1436	-	7470A	PPB	01 Oct 2018 14:11:23	U	SMPL
J5149-01LX5 - 1		317	0.0725	-	7470A	PPB	01 Oct 2018 14:13:30	U	SMPL
J5149-01A - 1		15275	4.1243	-	7470A	PPB	01 Oct 2018 14:15:38	U	SMPL
CCV40 - 1	CCV40	18617	5.0296	-	7470A	PPB	01 Oct 2018 14:17:45	U	SMPL
CCB40 - 1	CCB40	-176	-0.0611	-	7470A	PPB	01 Oct 2018 14:19:54	U	SMPL