

Analysis Method: SM2510 BANALYST: AMANDEEPParameter: ConductanceSUPERVISOR REVIEW BY: apatelRun Number: LB98257

Reagent/Standard	Lot/Log #
conductivity standard	WP67840

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB98257BL		1	22.5	0.90	0.95	10/01/2018	11:28
2	LB98257BS	1413	1	22.4	1340.00	1410.02	10/01/2018	11:35
3	J4880-07		1	22.9	948.00	987.61	10/01/2018	11:43

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--