

Analysis Method: 120.1

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB98258

Reagent/Standard	Lot/Log #
conductivity standard	WP67840

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB98258BL		1	22.5	0.90	0.95	10/01/2018	11:28
2	LB98258BS	1413	1	22.4	1340.00	1410.02	10/01/2018	11:35
3	J4880-08		1	22.9	946.00	985.53	10/01/2018	11:50

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$