

LB 918268


245.1

INSTRUMENT ID: CV 1

Linear

μ Abs.:

35146



A= 0.0000e+000

B= 2.8090e-004

C= -4.3379e-002

Rho= 0.9996103

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.045	-0.045	-4	0.000	-4				
0.05	0.050	0.042	-0.008	303	0.0 %	303				
0.2	0.200	0.172	-0.028	765	0.0 %	765				
2.5	2.500	2.466	-0.034	8935	0.0 %	8935				
5.0	5.000	5.168	0.168	18551	0.0 %	18551				
7.5	7.500	7.618	0.118	27274	0.0 %	27274				
10.0	10.000	9.829	-0.171	35146	0.0 %	35146				

LB98268 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 01 Oct 2018 15:02:35

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	SA	-4	-	0.0000245.1	PPB	01 Oct 2018 15:06:59	S	Std
0.05 - 1	SA-05	303	-	0.0500245.1	PPB	01 Oct 2018 15:09:17	S	Std
0.2 - 1	SA-2	765	-	0.2000245.1	PPB	01 Oct 2018 15:11:35	S	Std
2.5 - 1	SA-5	8935	-	2.5000245.1	PPB	01 Oct 2018 15:13:53	S	Std
5.0 - 1	SA-5	18551	-	5.0000245.1	PPB	01 Oct 2018 15:16:10	S	Std
7.5 - 1	SA-5	27274	-	7.5000245.1	PPB	01 Oct 2018 15:18:27	S	Std
10.0 - 1	SA-10	35146	-	10.0000245.1	PPB	01 Oct 2018 15:28:45	S	Std
ICV93 - 1	ICV93	14094	3.9157	-245.1	PPB	01 Oct 2018 15:31:36	U	SMPL
ICB93 - 1	ICB93	-30	-0.0518	-245.1	PPB	01 Oct 2018 15:33:52	U	SMPL
CCV41 - 1	CCV41	18965	5.2839	-245.1	PPB	01 Oct 2018 15:36:11	U	SMPL
CCB41 - 1	CCB41	-92	-0.0692	-245.1	PPB	01 Oct 2018 15:38:26	U	SMPL
CRA - 1	CRA	285	0.0367	-245.1	PPB	01 Oct 2018 15:40:45	U	SMPL
HighStd - 1	HighStd	35245	9.8570	-245.1	PPB	01 Oct 2018 15:43:00	U	SMPL
ChkStd - 1	ChkStd	24721	6.9008	-245.1	PPB	01 Oct 2018 15:45:16	U	SMPL
PB113347BL - 1	PBW	161	0.0018	-245.1	PPB	01 Oct 2018 15:47:36	U	SMPL
PB113347BS - 1	LCSW	14314	3.9775	-245.1	PPB	01 Oct 2018 15:49:55	U	SMPL
J5076-01 - 1	092418-SIDI-F-D-B	187	0.0091	-245.1	PPB	01 Oct 2018 15:52:13	U	SMPL
J5076-02 - 1	092418-SIDI-F-D-M	250	0.0268	-245.1	PPB	01 Oct 2018 15:54:33	U	SMPL
J5076-03 - 1	092418-SIDI-F-D-S	208	0.0150	-245.1	PPB	01 Oct 2018 15:56:50	U	SMPL
J5076-05 - 1	092418-SIDI-F-U-M	277	0.0344	-245.1	PPB	01 Oct 2018 16:01:24	U	SMPL
CCV42 - 1	CCV42	18458	5.1415	-245.1	PPB	01 Oct 2018 16:14:12	U	SMPL
CCB42 - 1	CCB42	-135	-0.0813	-245.1	PPB	01 Oct 2018 16:16:28	U	SMPL
J5076-06 - 1	092418-SIDI-F-U-S	183	0.0080	-245.1	PPB	01 Oct 2018 16:18:45	U	SMPL
J5076-07 - 1	092418-TIPC-F-D-B	237	0.0232	-245.1	PPB	01 Oct 2018 16:21:01	U	SMPL
J5076-08 - 1	092418-TIPC-F-D-S	159	0.0013	-245.1	PPB	01 Oct 2018 16:23:18	U	SMPL
J5076-09 - 1	092418-TIPC-F-U-B	157	0.0007	-245.1	PPB	01 Oct 2018 16:25:35	U	SMPL
J5076-11 - 1	092418-DBRI-E-D-B	241	0.0243	-245.1	PPB	01 Oct 2018 16:30:09	U	SMPL
J5076-12 - 1	092418-DBRI-E-D-S	267	0.0316	-245.1	PPB	01 Oct 2018 16:32:26	U	SMPL
J5076-13 - 1	092418-DBRI-E-U-B	183	0.0080	-245.1	PPB	01 Oct 2018 16:34:44	U	SMPL
J5076-14 - 1	092418-DBRI-E-U-S	251	0.0271	-245.1	PPB	01 Oct 2018 16:37:02	U	SMPL
CCV43 - 1	CCV43	18658	5.1977	-245.1	PPB	01 Oct 2018 16:41:35	U	SMPL
CCB43 - 1	CCB43	-58	-0.0597	-245.1	PPB	01 Oct 2018 16:43:52	U	SMPL
J5090-01 - 1	MW-1	11149	3.0884	-245.1	PPB	01 Oct 2018 16:50:42	U	SMPL
J5090-02 - 1	MW-2	34898	9.7595	-245.1	PPB	01 Oct 2018 16:52:59	U	SMPL
J5090-03 - 1	MW-4	2031	0.5271	-245.1	PPB	01 Oct 2018 16:55:16	U	SMPL
J5090-03DUP - 1	MW-4DUP	2251	0.5889	-245.1	PPB	01 Oct 2018 16:57:34	U	SMPL
J5090-03MS - 1	MW-4MS	17847	4.9699	-245.1	PPB	01 Oct 2018 16:59:51	U	SMPL
J5090-03MSD - 1	MW-4MSD	17591	4.8980	-245.1	PPB	01 Oct 2018 17:02:09	U	SMPL
PB113348BL - 1	PBW	11	-0.0403	-245.1	PPB	01 Oct 2018 17:04:26	U	SMPL
PB113348BS - 1	LCSW	14418	4.0067	-245.1	PPB	01 Oct 2018 17:06:43	U	SMPL
CCV44 - 1	CCV44	18672	5.2016	-245.1	PPB	01 Oct 2018 17:12:14	U	SMPL
CCB44 - 1	CCB44	-66	-0.0619	-245.1	PPB	01 Oct 2018 17:14:30	U	SMPL
J5133-01 - 1	092618-SIDI-F-D-B	151	-0.0010	-245.1	PPB	01 Oct 2018 17:16:48	U	SMPL
J5133-02 - 1	092618-SIDI-F-D-M	159	0.0013	-245.1	PPB	01 Oct 2018 17:19:05	U	SMPL
J5133-03 - 1	092618-SIDI-F-D-S	214	0.0167	-245.1	PPB	01 Oct 2018 17:21:21	U	SMPL
J5133-04 - 1	092618-SIDI-F-U-B	104	-0.0142	-245.1	PPB	01 Oct 2018 17:23:38	U	SMPL
J5133-05 - 1	092618-SIDI-F-U-M	103	-0.0144	-245.1	PPB	01 Oct 2018 17:25:54	U	SMPL
J5133-06 - 1	092618-SIDI-F-U-S	65	-0.0251	-245.1	PPB	01 Oct 2018 17:28:11	U	SMPL
J5133-07 - 1	092618-DBRI-F-D-B	101	-0.0150	-245.1	PPB	01 Oct 2018 17:30:28	U	SMPL
J5133-08 - 1	092618-DBRI-F-D-S	122	-0.0091	-245.1	PPB	01 Oct 2018 17:32:46	U	SMPL
J5133-09 - 1	092618-DBRI-F-U-B	172	0.0049	-245.1	PPB	01 Oct 2018 17:35:03	U	SMPL
J5133-10 - 1	092618-DBRI-F-U-S	204	0.0139	-245.1	PPB	01 Oct 2018 17:37:19	U	SMPL
CCV45 - 1	CCV45	18123	5.0474	-245.1	PPB	01 Oct 2018 17:39:35	U	SMPL
CCB45 - 1	CCB45	-86	-0.0675	-245.1	PPB	01 Oct 2018 17:41:52	U	SMPL
J5133-11 - 1	092618-TIPC-F-D-B	502	0.0976	-245.1	PPB	01 Oct 2018 17:44:10	U	SMPL
J5133-12 - 1	092618-TIPC-F-D-S	116	-0.0108	-245.1	PPB	01 Oct 2018 17:46:29	U	SMPL
J5133-13 - 1	092618-TIPC-F-U-B	234	0.0224	-245.1	PPB	01 Oct 2018 17:48:46	U	SMPL
J5133-14 - 1	092618-TIPC-F-U-S	168	0.0038	-245.1	PPB	01 Oct 2018 17:51:02	U	SMPL
J5133-16 - 1	092618-JETT-F-D-M	310	0.0437	-245.1	PPB	01 Oct 2018 17:55:36	U	SMPL
J5133-17 - 1	092618-JETT-F-D-S	261	0.0299	-245.1	PPB	01 Oct 2018 17:57:53	U	SMPL
J5133-18 - 1	092618-JETT-F-U-B	149	-0.0015	-245.1	PPB	01 Oct 2018 18:00:11	U	SMPL
J5133-19 - 1	092618-JETT-F-U-M	129	-0.0071	-245.1	PPB	01 Oct 2018 18:02:29	U	SMPL
J5133-15 - 1	092618-JETT-F-D-B	190	0.0100	-245.1	PPB	01 Oct 2018 18:05:20	U	SMPL
J5133-20 - 1	092618-JETT-F-U-S	263	0.0305	-245.1	PPB	01 Oct 2018 18:08:02	U	SMPL
CCV46 - 1	CCV46	17366	4.8348	-245.1	PPB	01 Oct 2018 18:12:55	U	SMPL
CCB46 - 1	CCB46	-98	-0.0709	-245.1	PPB	01 Oct 2018 18:15:11	U	SMPL
J5133-20DUP - 1	092618-JETT-F-U-SDUP	156	0.0004	-245.1	PPB	01 Oct 2018 18:17:29	U	SMPL
J5133-20MS - 1	092618-JETT-F-U-SMS	15631	4.3474	-245.1	PPB	01 Oct 2018 18:19:46	U	SMPL
J5133-20MSD - 1	092618-JETT-F-U-SMSD	15731	4.3755	-245.1	PPB	01 Oct 2018 18:22:02	U	SMPL
J5076-04 - 1	092418-SIDI-F-U-B	326	0.0482	-245.1	PPB	01 Oct 2018 18:24:19	U	SMPL
J5076-10 - 1	092418-TIPC-F-U-S	219	0.0181	-245.1	PPB	01 Oct 2018 18:26:36	U	SMPL
J5090-03LX5 - 1		671	0.1451	-245.1	PPB	01 Oct 2018 18:31:10	U	SMPL
J5090-03A - 1		18965	5.2839	-245.1	PPB	01 Oct 2018 18:33:28	U	SMPL
J5133-20LX5 - 1		-69	-0.0628	-245.1	PPB	01 Oct 2018 18:35:47	U	SMPL
J5133-20A - 1		16924	4.7106	-245.1	PPB	01 Oct 2018 18:38:06	U	SMPL
CCV47 - 1	CCV47	17016	4.7364	-245.1	PPB	01 Oct 2018 18:40:23	U	SMPL
CCB47 - 1	CCB47	-11	-0.0465	-245.1	PPB	01 Oct 2018 18:42:40	U	SMPL