



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98286

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	10/01/2018	16:50
2	CAL2	1	Water			21.9	7.00	10/01/2018	16:52
3	CAL3	1	Water			21.7	10.04	10/01/2018	16:54
4	ICV	1	Water			21.9	7.01	10/01/2018	16:56
5	CCV1	1	Water			21.8	2.00	10/01/2018	17:15
6	J5188-03	1	Water			16.2	9.83	10/01/2018	17:17
7	J5188-03DUP	1	Water			16.3	9.84	10/01/2018	17:19
8	J5188-04	1	Water			16.8	8.89	10/01/2018	17:21
9	CCV2	1	Water			21.7	12.00	10/01/2018	17:23

WORKLIST(Hardcopy Internal Chain)

VB 98286

WorkList Name : corrosivity100118-1

WorkList ID : 117410

Date : 10/1/2018 4:34:43 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/28/2018	Water	J5188-03	Corrosivity	Cool 4 deg C	ENVI77	N13	WWT-TK-03.04.06	09/28/2018	9040C
09/28/2018	Water	J5188-04	Corrosivity	Cool 4 deg C	ENVI77	N13	WWT-TK-01.02.05	09/28/2018	9040C

Date/Time 10/01/18 4:40
 Received by: SP
 Relinquished by: CP

Date/Time 10/01/18 5:30
 Received by: CP
 Relinquished by: SP