



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB98307

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.9	10/02/2018	12:40
2	J5188-03		1	7	>100.0	>212.0	>212.0	10/02/2018	13:05
3	J5188-03DUP		1	7	>100.0	>212.0	>212.0	10/02/2018	13:32
4	J5188-04		1	9	>100.0	>212.0	>212.0	10/02/2018	13:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 98/307

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp10218

WorkList ID : 117454

Date : 10/2/2018 11:54:19 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/05/2018	Water	J5188-03 C	Flash Point	Cool 4 deg C	ENV177	N13	WWT-TK-03.04.06	09/28/2018	1010A
10/05/2018	Water	J5188-04 D	Flash Point	Cool 4 deg C	ENV177	N13	WWT-TK-01.02.05	09/28/2018	1010A

Date/Time 10/2/18 12:05 PM
Received by: KS
Relinquished by: CP

Date/Time 10/2/18 14:05
Received by: CP
Relinquished by: KS