



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 10/01/2018

TEMP1 IN: 103 10/01/2018 07:00 TEMP OUT: 104 10/01/2018 08:10

TEMP2 IN: 104 10/01/2018 09:00 TEMP OUT: 104 10/01/2018 10:00

TEMP3 IN: 105 10/01/2018 11:00 TEMP OUT: 105 10/01/2018 12:35

TEMP4 IN: 105 10/01/2018 13:10 TEMP OUT: 105 10/01/2018 15:00

Run Number: LB98313

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB98313BL	LB98313BL	1.4000	1.4000	100	1.4000	1.4000	1.4000	0.0000	0
2	LB98313BS	LB98313BS	1.3945	1.3945	100	1.4445	1.4445	1.4445	0.0500	500
3	J5173-01	DSN-001A-092818	1.4151	1.4151	800	1.4193	1.4193	1.4193	0.0042	5.3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS100118

WorkList ID : 117474

Date : 10/1/2018 7:38:08 AM

XB9833

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/05/2018	Water	J5173-01	TSS	Cool 4 deg C	PKFM01	N12	DSN-001A-092818	09/28/2018	SM2540 D

FP-9823888

Date/Time 10/01/18 09:00
 Received by: 
 Relinquished by:

Date/Time 10/01/18 12:00
 Received by: 
 Relinquished by: AS