



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98317

Slope : 98.5

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	10/02/2018	16:25
2	CAL2	1	Water			22.8	7.01	10/02/2018	16:27
3	CAL3	1	Water			22.8	10.04	10/02/2018	16:29
4	ICV	1	Water			22.7	7.01	10/02/2018	16:31
5	CCV1	1	Water			22.8	2.00	10/02/2018	16:33
6	J5214-02	1	Water			22.1	7.21	10/02/2018	16:35
7	J5214-02DUP	1	Water			22.2	7.22	10/02/2018	16:37
8	J5214-05	1	Water			22.7	4.83	10/02/2018	16:39
9	CCV2	1	Water			22.6	12.00	10/02/2018	16:41

WORKLIST(Hardcopy Internal Chain)

LB 98317

WorkList Name : corrosivity100218-1

WorkList ID : 117469

Date : 10/2/2018 4:12:59 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2018	Water	J5214-02	Corrosivity	Cool 4 deg C	PSEG01	J31	N48963	10/02/2018	9040C
10/02/2018	Water	J5214-05	Corrosivity	Cool 4 deg C	PSEG01	J31	N48719	10/02/2018	9040C

Date/Time 10/02/18 4:15

Received by: SP

Relinquished by: SP

Date/Time 10/02/18 5:30

Received by: SP

Relinquished by: SP