



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98320

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	10/02/2018	16:25
2	CAL2	1	Water			22.8	7.01	10/02/2018	16:27
3	CAL3	1	Water			22.8	10.04	10/02/2018	16:29
4	ICV	1	Water			22.7	7.01	10/02/2018	16:31
5	CCV1	1	Water			22.6	2.00	10/02/2018	17:22
6	J5222-01	1	Solid	20.02	20	23.2	3.95	10/02/2018	17:24
7	J5222-01DUP	1	Solid	20.03	20	23.3	3.96	10/02/2018	17:26
8	CCV2	1	Water			22.8	12.00	10/02/2018	17:28

WORKLIST(Hardcopy Internal Chain)

1398320

WorkList Name : pH100218

WorkList ID : 117470

Date : 10/2/2018 4:13:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2018	Solid	J5222-01	E pH	Cool 4 deg C	AREC01	N31	SS-1	10/02/2018	9045D

Date/Time 10/02/18 5:00

Received by: SP

Relinquished by: CP

Date/Time 10/02/18 5:30

Received by: CP

Relinquished by: SP