


LB 98322
INSTRUMENT ID: CV2

7471B



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	257	0.000	257				
0.2	0.200	0.205	0.005	1151	0.0 %	1151				
2.5	2.500	2.476	-0.024	11512	0.0 %	11512				
5.0	5.000	5.045	0.045	23227	0.0 %	23227				
7.5	7.500	7.433	-0.067	34118	0.0 %	34118				
10.0	10.000	10.033	0.033	45977	0.0 %	45977				

LB98322 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Oct 2018 11:30:15

Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	257	-	0.00007471B	PPB	03 Oct 2018 11:31:58	S	Std
0.2 - 1	50.2	1151	-	0.20007471B	PPB	03 Oct 2018 11:34:16	S	Std
2.5 - 1	52.5	11512	-	2.50007471B	PPB	03 Oct 2018 11:36:34	S	Std
5.0 - 1	55	23227	-	5.00007471B	PPB	03 Oct 2018 11:38:51	S	Std
7.5 - 1	57.5	34118	-	7.50007471B	PPB	03 Oct 2018 11:41:09	S	Std
10.0 - 1	516	45977	-	10.00007471B	PPB	03 Oct 2018 11:43:27	S	Std
ICV98 - 1	ICV98	17676	3.8278	-7471B	PPB	03 Oct 2018 11:46:17	U	SMPL
ICB98 - 1	ICB98	13	-0.0450	-7471B	PPB	03 Oct 2018 11:48:32	U	SMPL
CCV67 - 1	CCV67	23344	5.0705	-7471B	PPB	03 Oct 2018 11:50:49	U	SMPL
CCB67 - 1	CCB67	-13	-0.0507	-7471B	PPB	03 Oct 2018 11:53:04	U	SMPL
CRA - 1	CRA	1166	0.2078	-7471B	PPB	03 Oct 2018 11:55:22	U	SMPL
HighStd - 1	HighStd	46433	10.1330	-7471B	PPB	03 Oct 2018 11:57:39	U	SMPL
ChkStd - 1	ChkStd	32110	6.9925	-7471B	PPB	03 Oct 2018 11:59:56	U	SMPL
PB113452BL - 1	PBS	142	-0.0167	-7471B	PPB	03 Oct 2018 12:05:36	U	SMPL
PB113452BS - 1	LCSS	15621	3.3772	-7471B	PPB	03 Oct 2018 12:07:52	U	SMPL
J5192-01 - 1	AU-05-100218-A	4888	1.0239	-7471B	PPB	03 Oct 2018 12:10:09	U	SMPL
J5192-03 - 1	AU-05-100218-B	3225	0.6593	-7471B	PPB	03 Oct 2018 12:12:27	U	SMPL
J5194-01 - 1	HR-06-100218-A	4738	0.9910	-7471B	PPB	03 Oct 2018 12:14:46	U	SMPL
J5194-03 - 1	HR-06-100218-B	3080	0.6275	-7471B	PPB	03 Oct 2018 12:17:04	U	SMPL
J5194-05 - 1	HR-06-100218-C	5421	1.1408	-7471B	PPB	03 Oct 2018 12:19:20	U	SMPL
CCV68 - 1	CCV68	23314	5.0639	-7471B	PPB	03 Oct 2018 12:21:36	U	SMPL
CCB68 - 1	CCB68	-19	-0.0520	-7471B	PPB	03 Oct 2018 12:23:52	U	SMPL
J5198-01 - 1	CL-01-100218-A	5244	1.1020	-7471B	PPB	03 Oct 2018 12:26:10	U	SMPL
J5198-03 - 1	CL-01-100218-B	5187	1.0895	-7471B	PPB	03 Oct 2018 12:28:26	U	SMPL
J5198-05 - 1	CL-01-100218-C	4333	0.9022	-7471B	PPB	03 Oct 2018 12:30:42	U	SMPL
J5198-07 - 1	CL-01-100218-D	3996	0.8283	-7471B	PPB	03 Oct 2018 12:32:58	U	SMPL
J5198-09 - 1	CL-01-100218-E	2404	0.4793	-7471B	PPB	03 Oct 2018 12:35:15	U	SMPL
J5212-01 - 1	CG-SOIL-1	494	0.0605	-7471B	PPB	03 Oct 2018 12:37:32	U	SMPL
J5212-05 - 1	CG-SOIL-2	341	0.0269	-7471B	PPB	03 Oct 2018 12:39:49	U	SMPL
J5218-01 - 1	280-356	767	0.1204	-7471B	PPB	03 Oct 2018 12:42:07	U	SMPL
J5218-03 - 1	229-211-RBR250129	54895	11.9883	-7471B	PPB	03 Oct 2018 12:44:24	U	SMPL
J5218-05 - 1	293-230	103478	22.6405	-7471B	PPB	03 Oct 2018 12:46:41	U	SMPL
CCV69 - 1	CCV69	22770	4.9447	-7471B	PPB	03 Oct 2018 12:49:00	U	SMPL
CCB69 - 1	CCB69	-25	-0.0533	-7471B	PPB	03 Oct 2018 12:51:20	U	SMPL
J5219-01 - 1	FK-BPM-01-E10-F10-E10A-F10A	941	0.1585	-7471B	PPB	03 Oct 2018 12:53:38	U	SMPL
J5219-03 - 1	FK-BPM-01-E9-F9-E9A-F9A	955	0.1616	-7471B	PPB	03 Oct 2018 12:55:54	U	SMPL
J5222-01 - 1	SS-1	1655	0.3151	-7471B	PPB	03 Oct 2018 12:58:52	U	SMPL
J5222-01DUP - 1	SS-1DUP	1651	0.3142	-7471B	PPB	03 Oct 2018 13:01:08	U	SMPL
J5222-01MS - 1	SS-1MS	18045	3.9087	-7471B	PPB	03 Oct 2018 13:03:25	U	SMPL
J5222-01MSD - 1	SS-1MSD	17653	3.8227	-7471B	PPB	03 Oct 2018 13:05:42	U	SMPL
J5222-01LX5 - 1		227	0.0020	-7471B	PPB	03 Oct 2018 13:08:00	U	SMPL
J5222-01A - 1		19526	4.2334	-7471B	PPB	03 Oct 2018 13:10:17	U	SMPL
J5218-03DLX2 - 1		27477	5.9767	-7471B	PPB	03 Oct 2018 13:12:35	U	SMPL
J5218-05DLX5 - 1		20793	4.5112	-7471B	PPB	03 Oct 2018 13:14:51	U	SMPL
CCV70 - 1	CCV70	23303	5.0615	-7471B	PPB	03 Oct 2018 13:17:09	U	SMPL
CCB70 - 1	CCB70	33	-0.0406	-7471B	PPB	03 Oct 2018 13:19:25	U	SMPL