



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB98323

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	10/03/2018	09:10
2	J5214-02		1	8	>100.0	>212.0	>212.0	10/03/2018	09:40
3	J5214-02DUP		1	8	>100.0	>212.0	>212.0	10/03/2018	10:10
4	J5214-05		1	9	>100.0	>212.0	>212.0	10/03/2018	10:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 98323

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp10318

WorkList ID : 117494

Date : 10/3/2018 8:25:47 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/05/2018	Water	J5214-02 C	Flash Point	Cool 4 deg C	PSEG01	J31	N48963	10/02/2018	1010A
10/05/2018	Water	J5214-05 C	Flash Point	Cool 4 deg C	PSEG01	J31	N48719	10/02/2018	1010A

Date/Time 10/3/18 8:30 AM

Received by: KS

Relinquished by: CP

Date/Time 10/3/18 10:40 AM

Received by: CP

Relinquished by: KS