

LB 98335


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33827



A= 0.0000e+000

B= 2.9833e-004

C= -1.3809e-002

Rho= 0.9998691

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	59	0.000	59				
0.05	0.050	0.047	-0.003	205	0.0 %	205				
0.2	0.200	0.208	0.008	744	0.0 %	744				
2.5	2.500	2.514	0.014	8474	0.0 %	8474				
5.0	5.000	5.032	0.032	16914	0.0 %	16914				
7.5	7.500	7.366	-0.134	24738	0.0 %	24738				
10.0	10.000	10.078	0.078	33827	0.0 %	33827				

LB98335 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 03 Oct 2018 14:44:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	59	-	0.0000	245.1	PPB	03 Oct 2018 14:58:05	S	Std
0.05 - 1	50.05	205	-	0.0500	245.1	PPB	03 Oct 2018 15:00:22	S	Std
0.2 - 1	50.2	744	-	0.2000	245.1	PPB	03 Oct 2018 15:02:40	S	Std
2.5 - 1	52.5	8474	-	2.5000	245.1	PPB	03 Oct 2018 15:04:58	S	Std
5.0 - 1	55	16914	-	5.0000	245.1	PPB	03 Oct 2018 15:07:15	S	Std
7.5 - 1	57.5	24738	-	7.5000	245.1	PPB	03 Oct 2018 15:09:31	S	Std
10.0 - 1	510	33827	-	10.0000	245.1	PPB	03 Oct 2018 15:11:49	S	Std
ICV99 - 1	ICV99	13903	4.1339	-245.1		PPB	03 Oct 2018 15:14:43	U	SMPL
ICB99 - 1	ICB99	-149	-0.0583	-245.1		PPB	03 Oct 2018 15:16:58	U	SMPL
CCV71 - 1	CCV71	17251	5.1327	-245.1		PPB	03 Oct 2018 15:19:16	U	SMPL
CCB71 - 1	CCB71	-93	-0.0416	-245.1		PPB	03 Oct 2018 15:21:32	U	SMPL
CRA - 1	CRA	246	0.0596	-245.1		PPB	03 Oct 2018 15:26:25	U	SMPL
HighStd - 1	HighStd	33738	10.0513	-245.1		PPB	03 Oct 2018 15:28:41	U	SMPL
ChkStd - 1	ChkStd	23714	7.0609	-245.1		PPB	03 Oct 2018 15:30:57	U	SMPL
PB113468BL - 1	PBW	-178	-0.0669	-245.1		PPB	03 Oct 2018 15:33:16	U	SMPL
PB113468BS - 1	LCSW	13118	3.8997	-245.1		PPB	03 Oct 2018 15:38:09	U	SMPL
J5221-01 - 1	100118-JETT-E-D-B	-70	-0.0347	-245.1		PPB	03 Oct 2018 15:45:23	U	SMPL
J5221-02 - 1	100118-JETT-E-D-S	-2	-0.0144	-245.1		PPB	03 Oct 2018 15:47:41	U	SMPL
J5221-03 - 1	100118-JETT-E-U-B	-19	-0.0195	-245.1		PPB	03 Oct 2018 15:49:59	U	SMPL
J5221-04 - 1	100118-JETT-E-U-S	69	0.0068	-245.1		PPB	03 Oct 2018 15:52:14	U	SMPL
CCV72 - 1	CCV72	17008	5.0602	-245.1		PPB	03 Oct 2018 15:54:31	U	SMPL
CCB72 - 1	CCB72	-96	-0.0424	-245.1		PPB	03 Oct 2018 15:56:46	U	SMPL
J5221-05 - 1	100118-DBRI-E-D-B	-2	-0.0144	-245.1		PPB	03 Oct 2018 15:59:03	U	SMPL
J5221-06 - 1	100118-DBRI-E-D-S	55	0.0026	-245.1		PPB	03 Oct 2018 16:01:20	U	SMPL
J5221-07 - 1	100118-DBRI-E-U-B	37	-0.0028	-245.1		PPB	03 Oct 2018 16:03:36	U	SMPL
J5221-08 - 1	100118-DBRI-E-U-S	114	0.0202	-245.1		PPB	03 Oct 2018 16:05:52	U	SMPL
J5221-09 - 1	100118-TIPC-E-D-M	194	0.0441	-245.1		PPB	03 Oct 2018 16:08:09	U	SMPL
J5221-10 - 1	100118-TIPC-E-U-M	170	0.0369	-245.1		PPB	03 Oct 2018 16:10:26	U	SMPL
J5221-11 - 1	100118-SIDI-E-D-B	117	0.0211	-245.1		PPB	03 Oct 2018 16:12:43	U	SMPL
J5221-12 - 1	100118-SIDI-E-D-M	117	0.0211	-245.1		PPB	03 Oct 2018 16:15:01	U	SMPL
J5221-13 - 1	100118-SIDI-E-D-S	-36	-0.0245	-245.1		PPB	03 Oct 2018 16:17:19	U	SMPL
J5221-14 - 1	100118-SIDI-E-U-B	129	0.0247	-245.1		PPB	03 Oct 2018 16:19:35	U	SMPL
CCV73 - 1	CCV73	16874	5.0203	-245.1		PPB	03 Oct 2018 16:21:53	U	SMPL
CCB73 - 1	CCB73	-149	-0.0583	-245.1		PPB	03 Oct 2018 16:24:09	U	SMPL
J5221-15 - 1	100118-SIDI-E-U-M	-27	-0.0219	-245.1		PPB	03 Oct 2018 16:26:26	U	SMPL
J5221-16 - 1	100118-SIDI-E-U-S	33	-0.0040	-245.1		PPB	03 Oct 2018 16:28:42	U	SMPL
J5221-16DUP - 1	100118-SIDI-E-U-SDUP	31	-0.0046	-245.1		PPB	03 Oct 2018 16:30:58	U	SMPL
J5221-16MS - 1	100118-SIDI-E-U-SMS	15050	4.4761	-245.1		PPB	03 Oct 2018 16:33:14	U	SMPL
J5221-16MSD - 1	100118-SIDI-E-U-SMSD	14834	4.4117	-245.1		PPB	03 Oct 2018 16:35:31	U	SMPL
J5186-05 - 1	092718-SIDI-F-U-M	-55	-0.0302	-245.1		PPB	03 Oct 2018 16:37:48	U	SMPL
J4878-08 - 1	PT-HG-WP	84708	25.2574	-245.1		PPB	03 Oct 2018 16:44:45	U	SMPL
J5221-16LX5 - 1		-61	-0.0320	-245.1		PPB	03 Oct 2018 16:49:54	U	SMPL
J5221-16A - 1		15807	4.7019	-245.1		PPB	03 Oct 2018 16:54:48	U	SMPL
CCV74 - 1	CCV74	17095	5.0862	-245.1		PPB	03 Oct 2018 16:57:04	U	SMPL
CCB74 - 1	CCB74	-226	-0.0812	-245.1		PPB	03 Oct 2018 16:59:20	U	SMPL
J4878-08DLX5 - 1	PT-HG-WP	16495	4.9072	-245.1		PPB	03 Oct 2018 17:06:13	U	SMPL
CCV75 - 1	CCV75	16919	5.0337	-245.1		PPB	03 Oct 2018 17:08:30	U	SMPL
CCB75 - 1	CCB75	-99	-0.0433	-245.1		PPB	03 Oct 2018 17:10:48	U	SMPL