

LB 98340


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36772



A= 0.0000e+000

B= 2.7233e-004

C= -3.2257e-002

Rho= 0.9997976

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	49	0.000	49				
0.2	0.200	0.165	-0.035	724	0.0 %	724				
2.5	2.500	2.499	-0.001	9296	0.0 %	9296				
5.0	5.000	5.156	0.156	19050	0.0 %	19050				
7.5	7.500	7.417	-0.083	27353	0.0 %	27353				
10.0	10.000	9.982	-0.018	36772	0.0 %	36772				

LB98340 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Oct 2018 17:24:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	49	-	0.0000	7470A	PPB	03 Oct 2018 17:25:02	S	Std
0.2 - 1	50.2	724	-	0.2000	7470A	PPB	03 Oct 2018 17:27:11	S	Std
2.5 - 1	52.5	9296	-	2.5000	7470A	PPB	03 Oct 2018 17:29:19	S	Std
5.0 - 1	55	19050	-	5.0000	7470A	PPB	03 Oct 2018 17:31:27	S	Std
7.5 - 1	57.5	27353	-	7.5000	7470A	PPB	03 Oct 2018 17:33:36	S	Std
10.0 - 1	510	36772	-	10.0000	7470A	PPB	03 Oct 2018 17:35:44	S	Std
ICV01 - 1	ICV01	14287	3.8586	-	7470A	PPB	03 Oct 2018 17:38:28	U	SMPL
ICB01 - 1	ICB101	-140	-0.0704	-	7470A	PPB	03 Oct 2018 17:40:34	U	SMPL
CCV76 - 1	CCV76	19070	5.1612	-	7470A	PPB	03 Oct 2018 17:42:42	U	SMPL
CCB76 - 1	CCB76	-186	-0.0829	-	7470A	PPB	03 Oct 2018 17:44:49	U	SMPL
CRA - 1	CRA	791	0.1832	-	7470A	PPB	03 Oct 2018 17:46:56	U	SMPL
HighStd - 1	HighStd	36004	9.7729	-	7470A	PPB	03 Oct 2018 17:49:03	U	SMPL
ChkStd - 1	ChkStd	25817	6.9986	-	7470A	PPB	03 Oct 2018 17:51:10	U	SMPL
PB113470BL - 1	PBW	-149	-0.0728	-	7470A	PPB	03 Oct 2018 17:53:19	U	SMPL
PB113470BS - 1	LCSW	12379	3.3390	-	7470A	PPB	03 Oct 2018 17:55:29	U	SMPL
J4879-08 - 1	PT-HG-WP	88996	24.2044	-	7470A	PPB	03 Oct 2018 17:57:37	U	SMPL
J5214-01 - 1	N48963	177	0.0159	-	7470A	PPB	03 Oct 2018 18:02:27	U	SMPL
J5214-04 - 1	N48719	416	0.0810	-	7470A	PPB	03 Oct 2018 18:09:11	U	SMPL
J5214-04DUP - 1	N48719DUP	474	0.0968	-	7470A	PPB	03 Oct 2018 18:11:19	U	SMPL
J5214-04MS - 1	N48719MS	11513	3.1031	-	7470A	PPB	03 Oct 2018 18:17:57	U	SMPL
CCV77 - 1	CCV77	18618	5.0381	-	7470A	PPB	03 Oct 2018 18:20:04	U	SMPL
CCB77 - 1	CCB77	-175	-0.0799	-	7470A	PPB	03 Oct 2018 18:22:12	U	SMPL
J5214-04MSD - 1	N48719MSD	11719	3.1592	-	7470A	PPB	03 Oct 2018 18:24:20	U	SMPL
J5214-04LX5 - 1		37	-0.0222	-	7470A	PPB	03 Oct 2018 18:26:27	U	SMPL
J5214-04A - 1		14299	3.8619	-	7470A	PPB	03 Oct 2018 18:28:35	U	SMPL
J4879-08DLX5 - 1	PT-HG-WP	17741	4.7992	-	7470A	PPB	03 Oct 2018 18:39:26	U	SMPL
CCV78 - 1	CCV78	19214	5.2004	-	7470A	PPB	03 Oct 2018 18:46:07	U	SMPL
CCB78 - 1	CCB78	-99	-0.0592	-	7470A	PPB	03 Oct 2018 18:48:15	U	SMPL