

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 10/04/2018

TEMP1 IN: 103 10/04/2018 07:00 TEMP OUT: 104 10/04/2018 08:15

TEMP2 IN: 104 10/04/2018 09:15 TEMP OUT: 104 10/04/2018 10:20

TEMP3 IN: 105 10/04/2018 12:00 TEMP OUT: 105 10/04/2018 13:40

TEMP4 IN: 105 10/04/2018 14:22 TEMP OUT: 105 10/04/2018 16:00

Run Number: LB98349

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB98349BL	LB98349BL	1.4012	1.4012	100	1.4012	1.4012	1.4012	0.0000	0
2	LB98349BS	LB98349BS	1.4000	1.4000	100	1.4501	1.4501	1.4501	0.0501	501
3	J5225-01	MW-1	1.3913	1.3913	900	1.3917	1.3917	1.3917	0.0004	0.4
4	J5225-02	MW-3	1.4095	1.4095	900	1.4096	1.4096	1.4096	0.0001	0.1
5	J5225-03	MW-4	1.3993	1.3993	900	1.3995	1.3995	1.3995	0.0002	0.2
6	J5225-04	MW-2	1.4002	1.4002	400	1.4051	1.4051	1.4051	0.0049	12.3
7	J5227-02	COMP	1.3968	1.3968	50	1.4025	1.4025	1.4025	0.0057	114
8	J5227-02DUP	COMPDUP	1.3938	1.3938	50	1.3996	1.3996	1.3996	0.0058	116
9	J5231-02	EFF-WW	1.3986	1.3986	200	1.4012	1.4012	1.4012	0.0026	13
10	J5233-02	OUTFALL001-181002	1.3818	1.3818	900	1.3824	1.3824	1.3824	0.0006	0.7
11	J5235-12	CLEARWELL	1.3889	1.3889	900	1.3933	1.3933	1.3933	0.0044	4.9
12	J5243-01	DSN001	1.3992	1.3992	400	1.4030	1.4030	1.4030	0.0038	9.5
13	J5244-04	OILY-WATER	1.3886	1.3886	200	1.3898	1.3898	1.3898	0.0012	6



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 10/04/2018

TEMP1 IN: 103 10/04/2018 07:00 TEMP OUT: 104 10/04/2018 08:15

TEMP2 IN: 104 10/04/2018 09:15 TEMP OUT: 104 10/04/2018 10:20

TEMP3 IN: 105 10/04/2018 12:00 TEMP OUT: 105 10/04/2018 13:40

TEMP4 IN: 105 10/04/2018 14:22 TEMP OUT: 105 10/04/2018 16:00

Run Number: LB98349

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS100418-2

WorkList ID : 117538

Date : 10/4/2018 7:42:24 AM

2098349

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/09/2018	Water	J5225-01 <i>FE</i>	TSS	Cool 4 deg C	LOCK01	N32	MW-1	10/02/2018	SM2540 D
10/09/2018	Water	J5225-02 <i>FE</i>	TSS	Cool 4 deg C	LOCK01	N32	MW-3	10/02/2018	SM2540 D
10/09/2018	Water	J5225-03 <i>CE</i>	TSS	Cool 4 deg C	LOCK01	N32	MW-4	10/02/2018	SM2540 D
10/09/2018	Water	J5225-04 <i>CE</i>	TSS	Cool 4 deg C	LOCK01	N32	MW-2	10/02/2018	SM2540 D
10/10/2018	Water	J5227-02 <i>B</i>	TSS	Cool 4 deg C	ARAM01	N23	COMP	10/03/2018	SM2540 D
10/10/2018	Water	J5231-02 <i>B</i>	TSS	Cool 4 deg C	ARDM01	N41	EFF-WW	10/03/2018	SM2540 D
10/09/2018	Water	J5233-02 <i>FE</i>	TSS	Cool 4 deg C	SHAW22	N31	OUTFALL001-181002	10/02/2018	SM2540 D
10/10/2018	Water	J5235-12 <i>CE</i>	TSS	Cool 4 deg C	CARP01	N51	CLEARWELL	10/03/2018	SM2540 D
10/10/2018	Water	J5243-01 <i>FE</i>	TSS	Cool 4 deg C	NAT104	N42	DSN001	10/03/2018	SM2540 D
10/08/2018	Water	J5244-04 <i>CE</i>	TSS	Cool 4 deg C	PSEG01	J61	OILY-WATER	10/03/2018	SM2540 D

2098349

Date/Time

Received by:

Relinquished by:

10/04/18, 08:00
AS
CS

Date/Time

Received by:

Relinquished by:

10/04/18, 12:00
AS
AS