



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98353

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.5	4.01	10/03/2018	12:10
2	CAL2	1	Water			22.4	7.00	10/03/2018	12:12
3	CAL3	1	Water			22.6	10.04	10/03/2018	12:14
4	ICV	1	Water			22.4	7.01	10/03/2018	12:16
5	CCV1	1	Water			22.6	2.00	10/03/2018	12:18
6	J5225-01	1	Water			16.4	5.60	10/03/2018	12:20
7	J5225-01DUP	1	Water			16.5	5.61	10/03/2018	12:22
8	J5225-02	1	Water			16.7	6.14	10/03/2018	12:24
9	J5225-03	1	Water			17.5	6.20	10/03/2018	12:26
10	J5225-04	1	Water			18.0	6.44	10/03/2018	12:28
11	CCV2	1	Water			22.2	12.00	10/03/2018	12:30

WORKLIST(Hardcopy Internal Chain)

LB 98353

WorkList Name : pH100318

WorkList ID : 117509

Date : 10/3/2018 11:51:15 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2018	Water	J5225-01	pH	Cool 4 deg C	LOCK01	N32	MW-1	10/02/2018	SM 4500-PH B
10/02/2018	Water	J5225-02	pH	Cool 4 deg C	LOCK01	N32	MW-3	10/02/2018	SM 4500-PH B
10/02/2018	Water	J5225-03	pH	Cool 4 deg C	LOCK01	N32	MW-4	10/02/2018	SM 4500-PH B
10/02/2018	Water	J5225-04	pH	Cool 4 deg C	LOCK01	N32	MW-2	10/02/2018	SM 4500-PH B

Date/Time 10/03/18 12:00

Received by: SP

Relinquished by: CP

Date/Time 10/03/18 12:35

Received by: CP

Relinquished by: SP