

LB 98376


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

35622



A= 0.0000e+000

B= 2.7884e-004

C= 6.2148e-003

Rho= 0.9999025

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	-3	0.000	-3				
0.2	0.200	0.213	0.013	741	0.0 %	741				
2.5	2.500	2.435	-0.065	8712	0.0 %	8712				
5.0	5.000	5.021	0.021	17984	0.0 %	17984				
7.5	7.500	7.586	0.086	27184	0.0 %	27184				
10.0	10.000	9.939	-0.061	35622	0.0 %	35622				

LB98356 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Oct 2018 10:02:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-3	-	0.0000	7470A	PPB	04 Oct 2018 10:06:25	S	Std
0.2 - 1	50.2	741	-	0.2000	7470A	PPB	04 Oct 2018 10:08:34	S	Std
2.5 - 1	52.5	8712	-	2.5000	7470A	PPB	04 Oct 2018 10:10:42	S	Std
5.0 - 1	55	17984	-	5.0000	7470A	PPB	04 Oct 2018 10:12:51	S	Std
7.5 - 1	57.5	27184	-	7.5000	7470A	PPB	04 Oct 2018 10:14:59	S	Std
10.0 - 1	510	35622	-	10.0000	7470A	PPB	04 Oct 2018 10:17:06	S	Std
ICV02 - 1	ICV02	14164	3.9557	-7470A	PPB	04 Oct 2018 10:19:44	U	SMPL	
ICB02 - 1	ICB02	-226	-0.0568	-7470A	PPB	04 Oct 2018 10:21:51	U	SMPL	
CCV79 - 1	CCV79	17913	5.0011	-7470A	PPB	04 Oct 2018 10:23:58	U	SMPL	
CCB79 - 1	CCB79	-139	-0.0325	-7470A	PPB	04 Oct 2018 10:26:05	U	SMPL	
CRA - 1	CRA	749	0.2151	-7470A	PPB	04 Oct 2018 10:28:12	U	SMPL	
HighStd - 1	HighStd	35528	9.9129	-7470A	PPB	04 Oct 2018 10:30:19	U	SMPL	
ChkStd - 1	ChkStd	25737	7.1828	-7470A	PPB	04 Oct 2018 10:32:26	U	SMPL	
PB113479BL - 1	PBW	-242	-0.0613	-7470A	PPB	04 Oct 2018 10:34:35	U	SMPL	
PB113479BS - 1	LCSW	12960	3.6200	-7470A	PPB	04 Oct 2018 10:36:47	U	SMPL	
PB113479TB - 1	PB113445TB	-89	-0.0186	-7470A	PPB	04 Oct 2018 10:38:54	U	SMPL	
J5192-02 - 1	AU-05-100218-A	311	0.0929	-7470A	PPB	04 Oct 2018 10:41:05	U	SMPL	
J5192-04 - 1	AU-05-100218-B	-64	-0.0116	-7470A	PPB	04 Oct 2018 10:43:13	U	SMPL	
J5194-02 - 1	HR-06-100218-A	92	0.0319	-7470A	PPB	04 Oct 2018 10:45:21	U	SMPL	
J5194-04 - 1	HR-06-100218-B	9	0.0087	-7470A	PPB	04 Oct 2018 10:47:28	U	SMPL	
CCV80 - 1	CCV80	17809	4.9721	-7470A	PPB	04 Oct 2018 10:49:35	U	SMPL	
CCB80 - 1	CCB80	-98	-0.0211	-7470A	PPB	04 Oct 2018 10:51:42	U	SMPL	
J5194-06 - 1	HR-06-100218-C	19	0.0115	-7470A	PPB	04 Oct 2018 10:53:49	U	SMPL	
J5198-02 - 1	CL-01-100218-A	25	0.0132	-7470A	PPB	04 Oct 2018 10:55:56	U	SMPL	
J5198-04 - 1	CL-01-100218-B	0	0.0062	-7470A	PPB	04 Oct 2018 10:58:02	U	SMPL	
J5198-06 - 1	CL-01-100218-C	34	0.0157	-7470A	PPB	04 Oct 2018 11:00:11	U	SMPL	
J5198-08 - 1	CL-01-100218-D	-34	-0.0033	-7470A	PPB	04 Oct 2018 11:02:19	U	SMPL	
J5198-10 - 1	CL-01-100218-E	2	0.0068	-7470A	PPB	04 Oct 2018 11:04:26	U	SMPL	
J5211-01 - 1	CB-1018-S-TCLP-GRID41-1	14	0.0101	-7470A	PPB	04 Oct 2018 11:06:34	U	SMPL	
J5211-02 - 1	CB-1018-S-TCLP-GRID41-2	43	0.0182	-7470A	PPB	04 Oct 2018 11:08:43	U	SMPL	
J5211-03 - 1	CB-1018-S-TCLP-GRID50-1	85	0.0299	-7470A	PPB	04 Oct 2018 11:10:51	U	SMPL	
J5212-04 - 1	CG-SOIL-1	26	0.0135	-7470A	PPB	04 Oct 2018 11:12:58	U	SMPL	
CCV81 - 1	CCV81	17889	4.9944	-7470A	PPB	04 Oct 2018 11:15:05	U	SMPL	
CCB81 - 1	CCB81	-96	-0.0206	-7470A	PPB	04 Oct 2018 11:17:12	U	SMPL	
J5212-08 - 1	CG-SOIL-2	-29	-0.0019	-7470A	PPB	04 Oct 2018 11:19:21	U	SMPL	
J5218-02 - 1	280-356	32	0.0151	-7470A	PPB	04 Oct 2018 11:21:28	U	SMPL	
J5218-04 - 1	229-211-RBR250129	-9	0.0037	-7470A	PPB	04 Oct 2018 11:23:35	U	SMPL	
J5218-06 - 1	293-230	70	0.0257	-7470A	PPB	04 Oct 2018 11:25:43	U	SMPL	
J5218-06DUP - 1	293-230DUP	2	0.0068	-7470A	PPB	04 Oct 2018 11:27:50	U	SMPL	
J5218-06MS - 1	293-230MS	12509	3.4943	-7470A	PPB	04 Oct 2018 11:29:58	U	SMPL	
J5218-06MSD - 1	293-230MSD	12463	3.4814	-7470A	PPB	04 Oct 2018 11:32:06	U	SMPL	
PB113480BL - 1	PBW	-158	-0.0378	-7470A	PPB	04 Oct 2018 11:34:15	U	SMPL	
PB113480BS - 1	LCSW	12541	3.5032	-7470A	PPB	04 Oct 2018 11:36:26	U	SMPL	
PB113480TB - 1	PB113451TB	-107	-0.0236	-7470A	PPB	04 Oct 2018 11:38:32	U	SMPL	
CCV82 - 1	CCV82	17923	5.0039	-7470A	PPB	04 Oct 2018 11:40:40	U	SMPL	
CCB82 - 1	CCB82	-157	-0.0376	-7470A	PPB	04 Oct 2018 11:42:47	U	SMPL	
J5214-02 - 1	N48963	111	0.0372	-7470A	PPB	04 Oct 2018 11:44:56	U	SMPL	
J5214-02DUP - 1	N48963DUP	69	0.0255	-7470A	PPB	04 Oct 2018 11:47:03	U	SMPL	
J5214-02MS - 1	N48963MS	12173	3.4006	-7470A	PPB	04 Oct 2018 11:49:10	U	SMPL	
J5214-02MSD - 1	N48963MSD	12402	3.4644	-7470A	PPB	04 Oct 2018 11:51:18	U	SMPL	
J5214-05 - 1	N48719	-73	-0.0141	-7470A	PPB	04 Oct 2018 11:53:25	U	SMPL	
J5218-06LX5 - 1		28	0.0140	-7470A	PPB	04 Oct 2018 11:55:34	U	SMPL	
J5218-06A - 1		14051	3.9242	-7470A	PPB	04 Oct 2018 12:00:33	U	SMPL	
J5214-02LX5 - 1		-105	-0.0231	-7470A	PPB	04 Oct 2018 12:02:40	U	SMPL	
J5214-02A - 1		14015	3.9142	-7470A	PPB	04 Oct 2018 12:04:52	U	SMPL	
CCV83 - 1	CCV83	17909	5.0000	-7470A	PPB	04 Oct 2018 12:06:59	U	SMPL	
CCB83 - 1	CCB83	-83	-0.0169	-7470A	PPB	04 Oct 2018 12:09:07	U	SMPL	