

21398359

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

10/4/2018 9:27

Test: Phenolics-mg/l

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.063	0.0	0.109	
ICB1	0.008	0.0	0.004	
CCV1	1.063	0.0	0.109	
CCB1	-0.000	0.0	0.003	
PB113484BL	0.002	0.0	0.003	
PB113484BS	1.065	0.0	0.109	
J5164-10	0.012	0.0	0.004	
J5164-11	0.095	0.0	0.013	
J5164-12	0.012	0.0	0.004	
CCV2	1.056	0.0	0.108	
CCB2	0.002	0.0	0.004	

N	11
Mean	0.398
SD	0.5271
CV%	132.45

AS

Aquakem v. 7.2AQ1

Results from time period:

Thu Oct 04 08:33:58 2018

Thu Oct 04 08:34:08 2018

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0448	mg/l	10/4/2018 7:28:51	
0.1PPM	A	Phenolics-	P	0.1026	mg/l	10/4/2018 7:28:52	
0.25PPM	A	Phenolics-	P	0.2519	mg/l	10/4/2018 7:28:53	
0.50PPM	A	Phenolics-	P	0.504	mg/l	10/4/2018 7:28:54	
1.0PPM	A	Phenolics-	P	0.9958	mg/l	10/4/2018 7:28:55	
2.0PPM	A	Phenolics-	P	2.0009	mg/l	10/4/2018 7:28:56	
ICV1	S	Phenolics-	P	1.0634	mg/l	10/4/2018 8:33:58	
ICB1	S	Phenolics-	P	0.0076	mg/l	10/4/2018 8:33:59	
CCV1	S	Phenolics-	P	1.0635	mg/l	10/4/2018 8:34:00	
CCB1	S	Phenolics-	P	-0.0004	mg/l	10/4/2018 8:34:01	
PB113484BL	S	Phenolics-	P	0.0015	mg/l	10/4/2018 8:34:02	
PB113484BS	S	Phenolics-	P	1.0651	mg/l	10/4/2018 8:34:03	
J5164-10	S	Phenolics-	P	0.0115	mg/l	10/4/2018 8:34:04	
J5164-11	S	Phenolics-	P	0.0955	mg/l	10/4/2018 8:34:05	
J5164-12	S	Phenolics-	P	0.0118	mg/l	10/4/2018 8:34:06	
CCV2	S	Phenolics-	P	1.0558	mg/l	10/4/2018 8:34:07	
CCB2	S	Phenolics-	P	0.0023	mg/l	10/4/2018 8:34:08	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : As

10/4/2018 7:30


Test Phenolics-

Accepted 10/4/2018 7:30

Factor 10.06
Bias 0.003

Coeff. of det. 0.999974

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0448	0.0500	
2	Phenol-2	0.014	0.1026	0.1000	
3	Phenol-2	0.028	0.2519	0.2500	
4	Phenol-2	0.053	0.5040	0.5000	
5	Phenol-2	0.102	0.9958	1.0000	
6	Phenol-2	0.202	2.0009	2.0000	