



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB98360

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	10/04/2018	10:05
2	J5244-05		1	9	>100.0	>212.0	>212.0	10/04/2018	10:42
3	J5244-05DUP		1	9	>100.0	>212.0	>212.0	10/04/2018	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB98360

WorkList Name : FP10418

WorkList ID : 117561

Date : 10/4/2018 9:25:57 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/08/2018	Water	J5244-05	C Flash Point	Cool 4 deg C	PSEG01	J61	OILY-WATER	10/03/2018	1010A

Date/Time 10/4/18 9:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 10/4/18 11:15 AM
 Received by: CP
 Relinquished by: KS