

LB 98361


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43940



A= 0.0000e+000

B= 2.2854e-004

C= -3.1625e-002

Rho= 0.9999920

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	51	0.000	51				
0.2	0.200	0.224	0.024	1118	0.0 %	1118				
2.5	2.500	2.501	0.001	11080	0.0 %	11080				
5.0	5.000	5.000	-0.000	22016	0.0 %	22016				
7.5	7.500	7.485	-0.015	32890	0.0 %	32890				
10.0	10.000	10.010	0.010	43940	0.0 %	43940				

LB98361 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Oct 2018 12:18:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	51	-	0.0000	7471B	PPB	04 Oct 2018 12:19:28	S	Std
0.2 - 1	50.2	1118	-	0.2000	7471B	PPB	04 Oct 2018 12:21:46	S	Std
2.5 - 1	52.5	11080	-	2.5000	7471B	PPB	04 Oct 2018 12:24:04	S	Std
5.0 - 1	55	22016	-	5.0000	7471B	PPB	04 Oct 2018 12:26:21	S	Std
7.5 - 1	57.5	32890	-	7.5000	7471B	PPB	04 Oct 2018 12:28:38	S	Std
10.0 - 1	510	43940	-	10.0000	7471B	PPB	04 Oct 2018 12:30:56	S	Std
ICV03 - 1	ICV03	17275	3.9164	-7471B	PPB	PPB	04 Oct 2018 12:33:44	U	SMPL
ICB03 - 1	ICB03	-167	-0.0698	-7471B	PPB	PPB	04 Oct 2018 12:36:02	U	SMPL
CCV84 - 1	CCV84	22092	5.0173	-7471B	PPB	PPB	04 Oct 2018 12:38:19	U	SMPL
CCB84 - 1	CCB84	-187	-0.0744	-7471B	PPB	PPB	04 Oct 2018 12:40:35	U	SMPL
CRA - 1	CRA	1104	0.2207	-7471B	PPB	PPB	04 Oct 2018 12:42:53	U	SMPL
HighStd - 1	HighStd	44020	10.0287	-7471B	PPB	PPB	04 Oct 2018 12:45:09	U	SMPL
ChkStd - 1	ChkStd	30780	7.0029	-7471B	PPB	PPB	04 Oct 2018 12:47:26	U	SMPL
PB113489BL - 1	PBS	-166	-0.0696	-7471B	PPB	PPB	04 Oct 2018 12:49:45	U	SMPL
PB113489BS - 1	LCSS	15049	3.4077	-7471B	PPB	PPB	04 Oct 2018 12:52:04	U	SMPL
J5150-01 - 1	SSSI-0926-S-10B	3511	0.7708	-7471B	PPB	PPB	04 Oct 2018 12:54:21	U	SMPL
J5199-01 - 1	OK-01-100318	2019	0.4298	-7471B	PPB	PPB	04 Oct 2018 12:56:40	U	SMPL
J5204-01 - 1	SU-04-100318-A	4183	0.9244	-7471B	PPB	PPB	04 Oct 2018 12:58:58	U	SMPL
J5204-03 - 1	SU-04-100318-B	4621	1.0245	-7471B	PPB	PPB	04 Oct 2018 13:01:16	U	SMPL
J5204-03DUP - 1	SU-04-100318-BDUP	4497	0.9961	-7471B	PPB	PPB	04 Oct 2018 13:03:32	U	SMPL
CCV85 - 1	CCV85	21896	4.9725	-7471B	PPB	PPB	04 Oct 2018 13:05:48	U	SMPL
CCB85 - 1	CCB85	-173	-0.0712	-7471B	PPB	PPB	04 Oct 2018 13:08:04	U	SMPL
J5204-03MS - 1	SU-04-100318-BMS	20709	4.7012	-7471B	PPB	PPB	04 Oct 2018 13:10:20	U	SMPL
J5204-03MSD - 1	SU-04-100318-BMSD	21192	4.8116	-7471B	PPB	PPB	04 Oct 2018 13:12:36	U	SMPL
J5223-03 - 1	SSSI-1001-S-DH-01	1164	0.2344	-7471B	PPB	PPB	04 Oct 2018 13:14:52	U	SMPL
J5223-04 - 1	SSSI-1001-S-DH-02	633	0.1130	-7471B	PPB	PPB	04 Oct 2018 13:17:10	U	SMPL
J5238-01 - 1	MT-T-5-8	4228	0.9346	-7471B	PPB	PPB	04 Oct 2018 13:19:27	U	SMPL
J5240-01 - 1	WJ-18-10-03-TSHG-12	1528	0.3176	-7471B	PPB	PPB	04 Oct 2018 13:21:44	U	SMPL
J5240-02 - 1	WJ-18-10-03-TSHG-13	1833	0.3873	-7471B	PPB	PPB	04 Oct 2018 13:24:01	U	SMPL
J5240-03 - 1	WJ-18-10-03-TSHG-14	2070	0.4415	-7471B	PPB	PPB	04 Oct 2018 13:26:19	U	SMPL
J5242-01 - 1	FK-BPM-01-E9-F9-E9A-F9A	1161	0.2337	-7471B	PPB	PPB	04 Oct 2018 13:28:37	U	SMPL
CCV86 - 1	CCV86	22421	5.0925	-7471B	PPB	PPB	04 Oct 2018 13:30:53	U	SMPL
CCB86 - 1	CCB86	-128	-0.0609	-7471B	PPB	PPB	04 Oct 2018 13:33:09	U	SMPL
J5204-03LX5 - 1		958	0.1873	-7471B	PPB	PPB	04 Oct 2018 13:35:26	U	SMPL
J5204-03A - 1		21805	4.9517	-7471B	PPB	PPB	04 Oct 2018 13:37:43	U	SMPL
CCV87 - 1	CCV87	21690	4.9254	-7471B	PPB	PPB	04 Oct 2018 13:40:00	U	SMPL
CCB87 - 1	CCB87	-66	-0.0467	-7471B	PPB	PPB	04 Oct 2018 13:42:18	U	SMPL