



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98370

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP68324
Turbidity std, 40NTU	WP68318
turbidity std 5 NTU	WP68321
turbidity std 20 NTU	WP68319
Turbidity std 1 NTU	WP68322
turbidity std 0 NTU	WP68323
turbidity std 10 NTU	WP68320

Intercept: 0.218

Slope: 0.939

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.233	10/03/2018	11:50
2	CAL2	1	1	0.996	10/03/2018	11:54
3	CAL3	5	1	4.920	10/03/2018	11:58
4	CAL4	10	1	9.920	10/03/2018	12:05
5	CAL5	20	1	18.800	10/03/2018	12:12
6	CAL6	40	1	37.800	10/03/2018	12:16



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98370

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.460	9.842	10/03/2018	12:22
2	ICB		1	0.227	0.010	10/03/2018	12:30
3	CCV1	10	1	9.580	9.970	10/03/2018	12:35
4	CCB1		1	0.229	0.012	10/03/2018	12:38
5	LB98370BL		1	0.242	0.026	10/03/2018	12:45
6	J5164-10		1	0.570	0.375	10/03/2018	12:49
7	J5164-11		1	2.320	2.239	10/03/2018	12:55
8	J5164-12		1	0.569	0.374	10/03/2018	13:04
9	J5225-01		1	0.802	0.622	10/03/2018	13:09
10	J5225-01DUP		1	0.796	0.616	10/03/2018	13:14
11	J5225-02		1	0.772	0.590	10/03/2018	13:20
12	J5225-03		1	0.459	0.257	10/03/2018	13:26
13	J5225-04		1	28.500	30.119	10/03/2018	13:32
14	CCV2	10	1	9.540	9.928	10/03/2018	13:38
15	CCB2		1	0.246	0.030	10/03/2018	13:44

L398370

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TURBIDITY100318

WorkList ID : 117505

Date : 10/3/2018 10:55:47 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/30/2018	Water	J5164-10	Turbidity	Cool 4 deg C	CHEM02	Wet Chem L	LOD-MDL-WATER-04-QT4-20	09/28/2018	SM2130 B
09/30/2018	Water	J5164-11	Turbidity	Cool 4 deg C	CHEM02	Wet Chem L	LOQ-WATER-05-QT4-2018	09/28/2018	SM2130 B
09/30/2018	Water	J5164-12	Turbidity	Cool 4 deg C	CHEM02	Wet Chem L	MDL-WATER-06-QT4-2018	09/28/2018	SM2130 B
10/04/2018	Water	J5225-01	Turbidity	Cool 4 deg C	LOCK01	N32	MW-1	10/02/2018	SM2130 B
10/04/2018	Water	J5225-02	Turbidity	Cool 4 deg C	LOCK01	N32	MW-3	10/02/2018	SM2130 B
10/04/2018	Water	J5225-03	Turbidity	Cool 4 deg C	LOCK01	N32	MW-4	10/02/2018	SM2130 B
10/04/2018	Water	J5225-04	Turbidity	Cool 4 deg C	LOCK01	N32	MW-2	10/02/2018	SM2130 B

Date/Time 10-3-18 11:05
 Received by: AB
 Relinquished by: CP

Date/Time 10-3-18 14:00
 Received by: CP
 Relinquished by: AE