



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98375

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP68366
Turbidity std, 80NTU	WP68356
Turbidity std, 40NTU	WP68357
turbidity std 5 NTU	WP68362
turbidity std 20 NTU	WP68358
Turbidity std 1 NTU	WP68364
turbidity std 0 NTU	WP68365
turbidity std 10 NTU	WP68360

Intercept: 1.804

Slope: 0.841

Regression: 0.9984

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.230	10/04/2018	13:33
2	CAL2	1	1	4.920	10/04/2018	13:38
3	CAL3	5	1	4.920	10/04/2018	13:42
4	CAL4	10	1	9.860	10/04/2018	13:49
5	CAL5	20	1	18.600	10/04/2018	13:55
6	CAL6	40	1	36.800	10/04/2018	14:02
7	CAL7	80	1	68.500	10/04/2018	14:08



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98375

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.460	9.103	10/04/2018	14:12
2	ICB		1	0.233	-1.868	10/04/2018	14:18
3	CCV1	10	1	9.600	9.270	10/04/2018	14:23
4	CCB1		1	0.241	-1.859	10/04/2018	14:30
5	LB98375BL		1	0.236	-1.864	10/04/2018	14:36
6	J5245-01		1	58.900	67.891	10/04/2018	14:42
7	J5245-01DUP		1	58.600	67.534	10/04/2018	14:49
8	CCV2	10	1	9.580	9.246	10/04/2018	14:55
9	CCB2		1	0.251	-1.847	10/04/2018	15:00

UB988375

WORKLIST(Hardcopy Internal Chain)

WorkList Name : turbidity10418

WorkList ID : 117582

Date : 10/4/2018 11:56:16 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/05/2018	Water	J5245-01	Turbidity	Cool 4 deg C	LIRO01	N21	UDS-CH4-01	10/03/2018	SM2130 B

Date/Time 10-4-18 12:16
Received by: AR
Relinquished by: CA

Date/Time 10-4-18 15:15
Received by: CA
Relinquished by: AA