



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98382

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.4	4.01	10/04/2018	17:20
2	CAL2	1	Water			22.6	7.00	10/04/2018	17:22
3	CAL3	1	Water			22.7	10.04	10/04/2018	17:24
4	ICV	1	Water			22.4	7.01	10/04/2018	17:26
5	CCV1	1	Water			22.5	2.00	10/04/2018	17:28
6	J5244-04	1	Water			21.7	6.67	10/04/2018	17:30
7	J5244-04DUP	1	Water			21.8	6.68	10/04/2018	17:32
8	J5248-01	1	Water			22.6	6.80	10/04/2018	17:34
9	J5262-02	1	Water			21.9	6.07	10/04/2018	17:36
10	CCV2	1	Water			22.5	12.00	10/04/2018	17:38

LB 98382

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH1000418

WorkList ID : 117602

Date : 10/4/2018 4:10:54 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/03/2018	Water	J5244-04	pH	Cool 4 deg C	PSEG01	J61	OILY-WATER	10/03/2018	9040C
10/04/2018	Water	J5248-01	pH	Cool 4 deg C	PSEG01	I51	N48971	10/04/2018	9040C
10/04/2018	Water	J5262-02	pH	Cool 4 deg C	PSEG01	J31	N48964-BS	10/04/2018	9040C

Date/Time 10/04/18 5:10
 Received by: SP
 Relinquished by: CP

Date/Time 10/04/18 5:40
 Received by: CP
 Relinquished by: SP