



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 10/04/2018

TEMP1 IN: 103 10/04/2018 14:00 TEMP OUT: 104 10/04/2018 15:10

TEMP2 IN: 104 10/04/2018 16:10 TEMP OUT: 104 10/05/2018 07:00

TEMP3 IN: 105 10/05/2018 08:30 TEMP OUT: 105 10/05/2018 10:00

TEMP4 IN: 105 10/05/2018 10:40 TEMP OUT: 105 10/05/2018 12:30

Run Number: LB98386

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB98386BL	LB98386BL	1.4012	1.4012	100	1.4012	1.4012	1.4012	0.0000	0
2	LB98386BS	LB98386BS	1.3952	1.3952	100	1.4460	1.4460	1.4460	0.0508	508
3	J5245-01	UDS-CH-4-01	1.3967	1.3967	400	1.4135	1.4135	1.4135	0.0168	42
4	J5248-01	N48971	1.3872	1.3872	500	1.3952	1.3952	1.3952	0.0080	16
5	J5253-01	001-WILLETTS-PT-BLVD (SEP)	1.3887	1.3887	300	1.4060	1.4060	1.4060	0.0173	57.7
6	J5260-01	SPDES-1	1.3940	1.3940	900	1.3942	1.3942	1.3942	0.0002	0.2
7	J5262-02	N48964-BS	1.3909	1.3909	100	1.4059	1.4059	1.4059	0.0150	150
8	J5262-02DUP	N48964-BSDUP	1.3980	1.3980	100	1.4124	1.4124	1.4124	0.0144	144
9	J5266-01	15220032-2	1.3905	1.3905	50	1.4097	1.4097	1.4097	0.0192	384

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS100518

WorkList ID : 117608

Date : 10/5/2018 7:31:59 AM

21398386

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/08/2018	Water	J5245-01	TSS	Cool 4 deg C	LIRO01	N21	UDS-CH4-01	10/03/2018	SM2540 D
10/09/2018	Water	J5248-01	TSS	Cool 4 deg C	PSEG01	I51	N48971	10/04/2018	SM2540 D
10/10/2018	Water	J5253-01	TSS	Cool 4 deg C	TULL01	N22	001-WILLETS-PT-BLVD(SEP)	10/03/2018	SM2540 D
10/05/2018	Water	J5260-01	TSS	Cool 4 deg C	SKAN06	N42	SPDES-1	10/04/2018	SM2540 D
10/09/2018	Water	J5262-02	TSS	Cool 4 deg C	PSEG01	J31	N48964-BS	10/04/2018	SM2540 D
10/11/2018	Water	J5266-01	TSS	Cool 4 deg C	PHAR02	N32	15220032-2	10/04/2018	SM2540 D

FP-4823888

Date/Time 10/05/18 10:08:00
 Received by: 
 Relinquished by: 

Date/Time 10/05/18 11:00
 Received by: 
 Relinquished by: 