



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98391

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.4	4.01	10/04/2018	17:20
2	CAL2	1	Water			22.6	7.00	10/04/2018	17:22
3	CAL3	1	Water			22.7	10.04	10/04/2018	17:24
4	ICV	1	Water			22.4	7.01	10/04/2018	17:26
5	CCV1	1	Water			22.5	2.00	10/04/2018	17:45
6	J5179-04	1	Solid	20.02	20	23.4	5.15	10/04/2018	17:47
7	J5179-04DUP	1	Solid	20.03	20	23.5	5.16	10/04/2018	17:49
8	J5179-08	1	Solid	20.02	20	23.5	5.60	10/04/2018	17:51
9	J5195-02	1	Solid	20.02	20	23.8	7.71	10/04/2018	17:53
10	J5205-02	1	Solid	20.04	20	24.1	8.30	10/04/2018	17:55
11	CCV2	1	Water			22.6	12.00	10/04/2018	17:57

WORKLIST(Hardcopy Internal Chain)

LB98391

WorkList Name : CORROSIONITY100418

WorkList ID : 117600

Date : 10/4/2018 4:10:32 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/04/2018	Solid	J5179-04	Corrosivity	Cool 4 deg C	PSEG01	J31	T2-GRAVEL	10/04/2018	9045D
10/04/2018	Solid	J5179-08	Corrosivity	Cool 4 deg C	PSEG01	J31	T1-GRAVEL	10/04/2018	9045D
10/04/2018	Solid	J5195-02	Corrosivity	Cool 4 deg C	PSEG05	J41	JC-03-100418	10/04/2018	9045D
10/04/2018	Solid	J5205-02	Corrosivity	Cool 4 deg C	PSEG05	J33	TR-05-100418	10/04/2018	9045D

Date/Time 10/04/18 5:10

Received by: SP

Relinquished by: CP

Date/Time 10/04/18 6:00

Received by: CP

Relinquished by: SP