



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98394

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	10/02/2018	16:25
2	CAL2	1	Water			22.8	7.00	10/02/2018	16:27
3	CAL3	1	Water			22.8	10.04	10/02/2018	16:29
4	ICV	1	Water			22.7	7.01	10/02/2018	16:31
5	CCV1	1	Water			22.6	2.00	10/02/2018	17:24
6	J5192-02	1	Solid	20.02	20	24.0	7.65	10/02/2018	17:26
7	J5192-02DUP	1	Solid	20.03	20	24.1	7.66	10/02/2018	17:28
8	J5192-04	1	Solid	20.02	20	24.0	7.06	10/02/2018	17:30
9	CCV2	1	Water			22.7	12.00	10/02/2018	17:32

LB 98394

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity-2

WorkList ID : 117639

Date : 10/2/2018 4:16:35 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2018	Solid	J5192-02	Corrosivity	Cool 4 deg C	PSEG05		AU-05-100218-A	10/02/2018	9045D
10/02/2018	Solid	J5192-04	Corrosivity	Cool 4 deg C	PSEG05		AU-05-100218-B	10/02/2018	9045D

Date/Time 10/02/18 4:20
 Received by: SP
 Relinquished by: SP

Date/Time 10/02/18 5:35
 Received by: CP
 Relinquished by: SP