

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

10/5/2018 13:27

Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.451	0.0	0.083	
ICB1	-0.520	0.0	0.003	
CCV1	240.487	0.0	0.205	
CCB1	-0.454	0.0	0.003	
PB113569BL	-0.465	0.0	0.003	
PB113569BS	96.362	0.0	0.084	
LOWPB113569	9.099	0.0	0.011	
HIGHPB113569	488.395	0.0	0.412	
J5179-01	-0.032	0.0	0.004	
J5179-01DUP	-0.099	0.0	0.003	
J5179-01MS	36.125	0.0	0.034	
J5179-01MSD	35.456	0.0	0.033	
J5179-05	0.246	0.0	0.004	
J5195-01	3.815	0.0	0.007	
CCV2	241.728	0.0	0.206	
CCB2	-0.028	0.0	0.004	
J5205-01	4.350	0.0	0.007	
J5255-02	0.962	0.0	0.004	
J5255-05	-0.303	0.0	0.003	
J5212-05	0.071	0.0	0.004	
J5218-05	239.337	0.0	0.204	
CCV3	247.886	0.0	0.211	
CCB3	-0.375	0.0	0.003	

N	23
Mean	75.543
SD	128.5138
CV%	170.12

Aquakem v. 7.2AQ1

Results from time period:

Fri Oct 05 12:50:48 2018

Fri Oct 05 13:04:10 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.7241	µg/l	10/5/2018 10:47:05	
5.0PPBCN	A	Total CN	P	4.0949	µg/l	10/5/2018 10:47:06	
10PPBCN	A	Total CN	P	9.7831	µg/l	10/5/2018 10:47:07	
50PPBCN	A	Total CN	P	51.2417	µg/l	10/5/2018 10:47:08	
100PPBCN	A	Total CN	P	100.3844	µg/l	10/5/2018 10:47:09	
250PPBCN	A	Total CN	P	250.8153	µg/l	10/5/2018 10:47:10	
500PPBCN	A	Total CN	P	499.4047	µg/l	10/5/2018 10:47:11	
ICV1	S	Total CN	P	95.4513	µg/l	10/5/2018 12:50:48	
ICB1	S	Total CN	P	-0.5199	µg/l	10/5/2018 12:50:49	
CCV1	S	Total CN	P	240.4872	µg/l	10/5/2018 12:50:50	
CCB1	S	Total CN	P	-0.4542	µg/l	10/5/2018 12:50:51	
PB113569BL	S	Total CN	P	-0.4649	µg/l	10/5/2018 12:50:52	
PB113569BS	S	Total CN	P	96.3622	µg/l	10/5/2018 12:50:53	
LOWPB113569	S	Total CN	P	9.0986	µg/l	10/5/2018 12:50:54	
HIGHPB113569	S	Total CN	P	488.3954	µg/l	10/5/2018 12:50:55	
J5179-01	S	Total CN	P	-0.0318	µg/l	10/5/2018 12:50:56	
J5179-01DUP	S	Total CN	P	-0.0991	µg/l	10/5/2018 12:50:57	
J5179-01MS	S	Total CN	P	36.125	µg/l	10/5/2018 12:50:58	
J5179-01MSD	S	Total CN	P	35.4557	µg/l	10/5/2018 12:58:21	
J5179-05	S	Total CN	P	0.2462	µg/l	10/5/2018 12:58:22	
J5195-01	S	Total CN	P	3.8147	µg/l	10/5/2018 12:58:23	
CCV2	S	Total CN	P	241.7282	µg/l	10/5/2018 12:58:24	
CCB2	S	Total CN	P	-0.0279	µg/l	10/5/2018 12:58:25	
J5205-01	S	Total CN	P	4.3497	µg/l	10/5/2018 12:58:26	
J5255-02	S	Total CN	P	0.9624	µg/l	10/5/2018 12:58:27	
J5255-05	S	Total CN	P	-0.3028	µg/l	10/5/2018 12:58:28	
J5212-05	S	Total CN	P	0.0711	µg/l	10/5/2018 12:58:29	
J5218-05	S	Total CN	P	239.3372	µg/l	10/5/2018 12:58:30	
CCV3	S	Total CN	P	247.8862	µg/l	10/5/2018 13:04:04	
CCB3	S	Total CN	P	-0.375	µg/l	10/5/2018 13:04:05	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

10/5/2018 11:05


Test Total CN

Accepted 10/5/2018 11:05

Factor 1194
Bias 0.004

Coeff. of det. 0.999980

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.7241	0.0000	
2	5.0PPBCN	0.007	4.0949	5.0000	
3	10PPBCN	0.012	9.7831	10.0000	
4	50PPBCN	0.046	51.2417	50.0000	
5	100PPBCN	0.088	100.3844	100.0000	
6	250PPBCN	0.214	250.8153	250.0000	
7	500PPBCN	0.422	499.4047	500.0000	