


ISM02.4

LB98406
INSTRUMENT ID: CV2

Linear

 μ Abs.:

42182



A= 0.0000e+000

B= 2.3901e-004

C= 1.1139e-002

Rho= 0.9998177

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.032	0.032	86	0.000	86					-
0.2	0.200	0.217	0.017	860	0.0 %	860					8
2.5	2.500	2.541	0.041	10583	0.0 %	10583					2
5.0	5.000	4.880	-0.120	20371	0.0 %	20371					-2
7.5	7.500	7.438	-0.062	31075	0.0 %	31075					-1
10.0	10.000	10.093	0.093	42182	0.0 %	42182					+1

no3
10/5/18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	86	-		0 ISM02.4	PPB	10/5/2018 13:35	Std
	0.2 S0.2	860	-		0.2 ISM02.4	PPB	10/5/2018 13:38	Std
	2.5 S2.5	10583	-		2.5 ISM02.4	PPB	10/5/2018 13:40	Std
	5 S5	20371	-		5 ISM02.4	PPB	10/5/2018 13:42	Std
	7.5 S7.5	31075	-		7.5 ISM02.4	PPB	10/5/2018 13:45	Std
	10 S10.0	42182	-		10 ISM02.4	PPB	10/5/2018 13:47	Std
ICV	ICV	16023	3.8407 -		ISM02.4	PPB	10/5/2018 13:50	SMPL
ICB	ICB	-113	-0.0159 -		ISM02.4	PPB	10/5/2018 13:52	SMPL
CCV097	CCV097	20564	4.9261 -		ISM02.4	PPB	10/5/2018 13:54	SMPL
CCB097	CCB097	-134	-0.0209 -		ISM02.4	PPB	10/5/2018 13:56	SMPL
PB113545BL	PBW003	36	0.0197 -		ISM02.4	PPB	10/5/2018 13:59	SMPL
J5249-01	MC0AF6	46	0.0221 -		ISM02.4	PPB	10/5/2018 14:01	SMPL
J5249-02	MC0AF7	7	0.0128 -		ISM02.4	PPB	10/5/2018 14:03	SMPL
J5249-03	MC0AF8	29	0.0181 -		ISM02.4	PPB	10/5/2018 14:05	SMPL
J5249-04	MC0AF9	107	0.0367 -		ISM02.4	PPB	10/5/2018 14:08	SMPL
J5249-05	MC0AGO	21	0.0162 -		ISM02.4	PPB	10/5/2018 14:10	SMPL
J5249-06	MC0AG1	18	0.0154 -		ISM02.4	PPB	10/5/2018 14:12	SMPL
J5249-07	MC0AF6D	44	0.0217 -		ISM02.4	PPB	10/5/2018 14:15	SMPL
J5249-08	MC0AF6S	4171	1.008 -		ISM02.4	PPB	10/5/2018 14:17	SMPL
CCV098	CCV098	20079	4.8101 -		ISM02.4	PPB	10/5/2018 14:19	SMPL
CCB098	CCB098	-140	-0.0223 -		ISM02.4	PPB	10/5/2018 14:21	SMPL