

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB98417

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP68423
calibration std. phosphate 0.5 ppm	WP68424
calibration std. phosphate 0.3 ppm	WP68425
calibration std. 0 ppm	WP68430
phosphate ICV std.	WP68433
phosphate CCV std.	WP68434
calibration std. phosphate 0.1 ppm	WP68427
calibration std. phosphate 0.05 ppm	WP68429
Combined reagent	WP68432
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
Sodium hydroxide, 1N	WP66924

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	10/05/2018	12:39
2	CAL2	0.05	1	50	50	0.033	10/05/2018	12:39
3	CAL3	0.10	1	50	50	0.065	10/05/2018	12:40
4	CAL4	0.30	1	50	50	0.186	10/05/2018	12:40
5	CAL5	0.50	1	50	50	0.330	10/05/2018	12:41
6	CAL6	1.00	1	50	50	0.649	10/05/2018	12:41



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB98417

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.319	0.492	10/05/2018	12:44
2	ICB		1	50	50	0.000	0.001	10/05/2018	12:44
3	CCV1	0.50	1	50	50	0.325	0.501	10/05/2018	12:45
4	CCB1		1	50	50	0.000	0.001	10/05/2018	12:45
5	PB113581BL		1	50	50	0.000	0.001	10/05/2018	12:46
6	PB113581BS	0.50	1	50	50	0.318	0.490	10/05/2018	12:46
7	J5164-07		1	50	50	0.021	0.034	10/05/2018	12:47
8	J5164-08		1	50	50	0.091	0.141	10/05/2018	12:47
9	J5164-09		1	50	50	0.019	0.031	10/05/2018	12:48
10	J5190-01		1	50	50	1.130	1.739	10/05/2018	12:48
11	J5235-13		1	50	50	0.109	0.169	10/05/2018	12:49
12	J5235-13DUP		1	50	50	0.108	0.167	10/05/2018	12:49
13	J5235-13MS	0.50	1	50	50	0.430	0.663	10/05/2018	12:50
14	J5235-13MSD	0.50	1	50	50	0.431	0.664	10/05/2018	12:50
15	CCV2	0.50	1	50	50	0.328	0.506	10/05/2018	12:51
16	CCB2		1	50	50	0.000	0.001	10/05/2018	12:51
17	J5243-01		1	50	50	0.050	0.078	10/05/2018	12:52
18	J5190-01		5	50	50	0.291	0.449	10/05/2018	12:59
19	CCV3	0.50	1	50	50	0.326	0.503	10/05/2018	13:00
20	CCB3		1	50	50	0.000	0.001	10/05/2018	13:00