



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By :

Run Number: LB98419

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	10/05/2018	12:40
2	CAL2	1	Water			22.4	7.00	10/05/2018	12:42
3	CAL3	1	Water			22.2	10.04	10/05/2018	12:44
4	ICV	1	Water			22.1	7.01	10/05/2018	12:46
5	CCV1	1	Water			22.3	2.00	10/05/2018	12:48
6	J5301-01	1	Water			13.3	4.26	10/05/2018	12:50
7	J5301-01DUP	1	Water			13.4	4.27	10/05/2018	12:52
8	J5301-02	1	Water			12.6	4.38	10/05/2018	12:54
9	J5301-03	1	Water			15.2	5.96	10/05/2018	12:56
10	CCV2	1	Water			22.4	12.00	10/05/2018	12:58

B98419

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH100518 WorkList ID : 117646 Date : 10/5/2018 12:00:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/04/2018	Water	J5301-01	pH	Cool 4 deg C	STER09	O41	CELL-A-POOL WATER(DE-IO	10/04/2018	SM 4500-PH B
10/04/2018	Water	J5301-02	pH	Cool 4 deg C	STER09	O41	CELL-B-C-POOL WATER(DE-	10/04/2018	SM 4500-PH B
10/04/2018	Water	J5301-03	pH	Cool 4 deg C	STER09	O41	CELL-D-POOL WATER(DE-IO	10/04/2018	SM 4500-PH B

Date/Time 10/05/18 12:30
Received by: SP
Relinquished by: CP

Date/Time 10/05/18 1:10
Received by: CP
Relinquished by: SP