



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98422

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	10/05/2018	16:30
2	CAL2	1	Water			22.4	7.00	10/05/2018	16:32
3	CAL3	1	Water			22.8	10.04	10/05/2018	16:34
4	ICV	1	Water			22.6	7.01	10/05/2018	16:36
5	CCV1	1	Water			22.7	2.00	10/05/2018	16:38
6	J5304-01	1	Water			19.2	6.97	10/05/2018	16:40
7	J5304-01DUP	1	Water			19.3	6.98	10/05/2018	16:42
8	CCV2	1	Water			22.6	12.00	10/05/2018	16:44

WORKLIST(Hardcopy Internal Chain)

LB 96422

WorkList Name : corrosivity100518-1

WorkList ID : 117659

Date : 10/5/2018 4:16:03 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/04/2018	Water	J5304-01	Corrosivity	Cool 4 deg C	TETR06	N31	CA-AQIDW-20181004	10/04/2018	9040C

Date/Time 10/05/18 4:30
 Received by: SP
 Relinquished by: CP

Date/Time 10/05/18 5:00
 Received by: CP
 Relinquished by: SP