


7471B

LB98423
INSTRUMENT ID : LV1

Linear

 μ Abs.:

4280E



A= 0.0000e+000

B= 2.3407e-004

C= 8.3881e-003

Rho= 0.9999569

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	29	0.000	29				
0.2	0.200	0.233	0.033	960	0.0 %	960				
2.5	2.500	2.437	-0.063	10376	0.0 %	10376				
5.0	5.000	5.014	0.014	21387	0.0 %	21387				
7.5	7.500	7.472	-0.028	31888	0.0 %	31888				
10.0	10.000	10.028	0.028	42806	0.0 %	42806				

LB98423 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Oct 2018 17:02:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	29	-	0.00007471B		PPB	05 Oct 2018 17:05:21	S	Std
0.2 - 1	50.2	960	-	0.20007471B		PPB	05 Oct 2018 17:07:38	S	Std
2.5 - 1	52.5	10376	-	2.50007471B		PPB	05 Oct 2018 17:09:56	S	Std
5.0 - 1	55	21387	-	5.00007471B		PPB	05 Oct 2018 17:12:13	S	Std
7.5 - 1	57.5	31888	-	7.50007471B		PPB	05 Oct 2018 17:14:31	S	Std
10.0 - 1	810	42806	-	10.00007471B		PPB	05 Oct 2018 17:19:49	S	Std
ICV06 - 1	ICV06	17456	4.0943	-7471B		PPB	05 Oct 2018 17:22:44	U	SMPL
ICB06 - 1	ICB06	-133	-0.0227	-7471B		PPB	05 Oct 2018 17:24:59	U	SMPL
CCV08 - 1	CCV08	21642	5.0741	-7471B		PPB	05 Oct 2018 17:27:15	U	SMPL
CCB08 - 1	CCB08	-77	-0.0096	-7471B		PPB	05 Oct 2018 17:29:32	U	SMPL
CRA - 1	CRA	873	0.2127	-7471B		PPB	05 Oct 2018 17:31:50	U	SMPL
HighStd - 1	HighStd	43147	10.1077	-7471B		PPB	05 Oct 2018 17:34:06	U	SMPL
ChkStd - 1	ChkStd	29414	6.8933	-7471B		PPB	05 Oct 2018 17:36:22	U	SMPL
PB113541BL - 1	PBS	-204	-0.0394	-7471B		PPB	05 Oct 2018 17:38:41	U	SMPL
PB113541BS - 1	LCSS	16617	3.8979	-7471B		PPB	05 Oct 2018 17:40:59	U	SMPL
J5179-01 - 1	T2-GRAVEL	958	0.2326	-7471B		PPB	05 Oct 2018 17:43:16	U	SMPL
J5179-05 - 1	T1-GRAVEL	801	0.1959	-7471B		PPB	05 Oct 2018 17:45:34	U	SMPL
J5195-01 - 1	JC-02-100418	9660	2.2695	-7471B		PPB	05 Oct 2018 17:47:52	U	SMPL
J5205-01 - 1	TR-05-100418	1069	0.2586	-7471B		PPB	05 Oct 2018 17:50:10	U	SMPL
J5255-02 - 1	TRACK-D-1	1129	0.2727	-7471B		PPB	05 Oct 2018 17:52:25	U	SMPL
CCV09 - 1	CCV09	21637	5.0729	-7471B		PPB	05 Oct 2018 17:54:41	U	SMPL
CCB09 - 1	CCB09	-52	-0.0038	-7471B		PPB	05 Oct 2018 17:56:57	U	SMPL
J5255-05 - 1	TRACK-D-2	1167	0.2815	-7471B		PPB	05 Oct 2018 17:59:13	U	SMPL
J5255-05DUP - 1	TRACK-D-2DUP	1128	0.2724	-7471B		PPB	05 Oct 2018 18:01:29	U	SMPL
J5255-05MS - 1	TRACK-D-2MS	19514	4.5760	-7471B		PPB	05 Oct 2018 18:03:44	U	SMPL
J5255-05MSD - 1	TRACK-D-2MSD	17834	4.1828	-7471B		PPB	05 Oct 2018 18:06:01	U	SMPL
J5263-01 - 1	GRID10-COMP-100318	7305	1.7183	-7471B		PPB	05 Oct 2018 18:08:18	U	SMPL
J5263-02 - 1	GRID11-COMP-100318	62686	14.6812	-7471B		PPB	05 Oct 2018 18:10:35	U	SMPL
J5263-03 - 1	GRID12-COMP-100318	59091	13.8397	-7471B		PPB	05 Oct 2018 18:12:52	U	SMPL
J5263-04 - 1	GRID13-COMP-100318	23139	5.4245	-7471B		PPB	05 Oct 2018 18:15:13	U	SMPL
J5263-05 - 1	GRID14-COMP-100318	3152	0.7462	-7471B		PPB	05 Oct 2018 18:17:33	U	SMPL
J5263-06 - 1	GRID15-COMP-100318	21510	5.0432	-7471B		PPB	05 Oct 2018 18:19:50	U	SMPL
CCV10 - 1	CCV10	21463	5.0322	-7471B		PPB	05 Oct 2018 18:22:06	U	SMPL
CCB10 - 1	CCB10	-127	-0.0213	-7471B		PPB	05 Oct 2018 18:24:24	U	SMPL
J5263-07 - 1	GRID16-COMP-100318	4983	1.1747	-7471B		PPB	05 Oct 2018 18:26:42	U	SMPL
J5263-08 - 1	GRID17-COMP-100318	7227	1.7000	-7471B		PPB	05 Oct 2018 18:28:58	U	SMPL
J5263-09 - 1	GRID18-COMP-100318	3275	0.7750	-7471B		PPB	05 Oct 2018 18:31:15	U	SMPL
J5263-10 - 1	GRID19-COMP-100318	2259	0.5371	-7471B		PPB	05 Oct 2018 18:33:31	U	SMPL
J5255-05LX5 - 1		210	0.0575	-7471B		PPB	05 Oct 2018 18:35:48	U	SMPL
J5255-05A - 1		18749	4.3969	-7471B		PPB	05 Oct 2018 18:38:05	U	SMPL
J5263-02DLX2 - 1		30871	7.2343	-7471B		PPB	05 Oct 2018 18:40:23	U	SMPL
J5263-03DLX2 - 1		28985	6.7928	-7471B		PPB	05 Oct 2018 18:42:41	U	SMPL
CCV11 - 1	CCV11	21972	5.1513	-7471B		PPB	05 Oct 2018 18:45:01	U	SMPL
CCB11 - 1	CCB11	-129	-0.0218	-7471B		PPB	05 Oct 2018 18:47:18	U	SMPL