



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98430

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	10/05/2018	16:30
2	CAL2	1	Water			22.4	7.00	10/05/2018	16:32
3	CAL3	1	Water			22.8	10.04	10/05/2018	16:34
4	ICV	1	Water			22.6	7.01	10/05/2018	16:36
5	CCV1	1	Water			22.4	2.00	10/05/2018	17:00
6	J5270-08	1	Solid	20.02	20	23.1	8.56	10/05/2018	17:02
7	J5270-08DUP	1	Solid	20.03	20	23.2	8.57	10/05/2018	17:04
8	J5300-04	1	Solid	20.03	20	22.2	11.21	10/05/2018	17:06
9	J5305-03	1	Solid	20.02	20	23.4	6.83	10/05/2018	17:08
10	J5307-01	1	Solid	20.04	20	23.3	5.12	10/05/2018	17:10
11	CCV2	1	Water			22.6	12.00	10/05/2018	17:12

WORKLIST(Hardcopy Internal Chain)

LB94430

WorkList Name : corrosivity100518

WorkList ID : 117658

Date : 10/5/2018 3:59:35 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/05/2018	Solid	J5270-08	Corrosivity	Cool 4 deg C	PSEG01	P11	PHASE-3-SOIL	10/05/2018	9045D
10/05/2018	Solid	J5300-04	Corrosivity	Cool 4 deg C	PSEG01	J61	SR-STONE	10/05/2018	9045D
10/05/2018	Solid	J5305-03	Corrosivity	Cool 4 deg C	PSEG01	P11	331-SAND-SOIL	10/05/2018	9045D
10/05/2018	Solid	J5307-01	Corrosivity	Cool 4 deg C	PSEG01	P11	148-20-OLD	10/05/2018	9045D

Date/Time 10/05/18 4:20

Received by: SP

Relinquished by: CP

Date/Time 10/05/18 5:20

Received by: CP

Relinquished by: SP