



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB98435

Ambient Barometric Pressure (mmHg): 765

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.9	10/08/2018	09:20
2	J5304-01		1	7	>100.0	>212.0	>212.0	10/08/2018	09:50
3	J5304-01DUP		1	7	>100.0	>212.0	>212.0	10/08/2018	10:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB98435

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp10818

WorkList ID : 117688

Date : 10/8/2018 8:43:48 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/18/2018	Water	J5304-01	Flash Point	Cool 4 deg C	TETR06	N31	CA-AQIDW-20181004	10/04/2018	1010A

Date/Time 10/8/18 8:45 AM
Received by: KS
Relinquished by: CP

Date/Time 10/8/18 10:30 AM
Received by: CP
Relinquished by: KS