

LB98440


7471B

INSTRUMENT ID: CV4

Linear

 μ Abs.:

43471



A= 0.0000e+000

B= 2.2851e-004

C= -2.6469e-002

Rho= 0.9995338

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.013	-0.013	60	0.000	60				
0.2	0.200	0.178	-0.022	894	0.0 %	894				
2.5	2.500	2.597	0.097	11480	0.0 %	11480				
5.0	5.000	4.847	-0.153	21327	0.0 %	21327				
7.5	7.500	7.684	0.184	33745	0.0 %	33745				
10.0	10.000	9.907	-0.093	43471	0.0 %	43471				

LB98440 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Oct 2018 13:53:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	60	-	0.0000	7471B	PPB	08 Oct 2018 13:55:34	S	Std
0.2 - 1	50.2	894	-	0.2000	7471B	PPB	08 Oct 2018 13:57:52	S	Std
2.5 - 1	52.5	11480	-	2.5000	7471B	PPB	08 Oct 2018 14:00:10	S	Std
5.0 - 1	55	21327	-	5.0000	7471B	PPB	08 Oct 2018 14:02:28	S	Std
7.5 - 1	57.5	33745	-	7.5000	7471B	PPB	08 Oct 2018 14:04:45	S	Std
10.0 - 1	510	43471	-	10.0000	7471B	PPB	08 Oct 2018 14:07:02	S	Std
ICV11 - 1	ICV11	17359	3.9402	-	7471B	PPB	08 Oct 2018 14:09:58	U	SMPL
ICB11 - 1	ICB11	-98	-0.0489	-	7471B	PPB	08 Oct 2018 14:12:14	U	SMPL
CCV17 - 1	CCV17	22544	5.1250	-	7471B	PPB	08 Oct 2018 14:14:30	U	SMPL
CCB17 - 1	CCB17	-198	-0.0717	-	7471B	PPB	08 Oct 2018 14:16:46	U	SMPL
CRA - 1	CRA	820	0.1609	-	7471B	PPB	08 Oct 2018 14:19:04	U	SMPL
HighStd - 1	HighStd	43254	9.8573	-	7471B	PPB	08 Oct 2018 14:21:20	U	SMPL
ChkStd - 1	ChkStd	30685	6.9852	-	7471B	PPB	08 Oct 2018 14:23:36	U	SMPL
PB113585BL - 1	PBS	-134	-0.0571	-	7471B	PPB	08 Oct 2018 14:25:55	U	SMPL
PB113585BS - 1	LCSS	16011	3.6321	-	7471B	PPB	08 Oct 2018 14:28:13	U	SMPL
J5270-07 - 1	PHASE-3-SOIL	6534	1.4666	-	7471B	PPB	08 Oct 2018 14:30:29	U	SMPL
J5287-02 - 1	TRACK-D-3	629	0.1173	-	7471B	PPB	08 Oct 2018 14:32:47	U	SMPL
J5287-02DUP - 1	TRACK-D-3DUP	638	0.1193	-	7471B	PPB	08 Oct 2018 14:35:04	U	SMPL
J5287-02MS - 1	TRACK-D-3MS	17544	3.9824	-	7471B	PPB	08 Oct 2018 14:37:22	U	SMPL
J5287-02MSD - 1	TRACK-D-3MSD	17026	3.8641	-	7471B	PPB	08 Oct 2018 14:39:38	U	SMPL
CCV18 - 1	CCV18	21536	4.8946	-	7471B	PPB	08 Oct 2018 14:41:54	U	SMPL
CCB18 - 1	CCB18	-147	-0.0601	-	7471B	PPB	08 Oct 2018 14:44:11	U	SMPL
J5287-05 - 1	TRACK-D-4	1118	0.2290	-	7471B	PPB	08 Oct 2018 14:46:28	U	SMPL
J5287-08 - 1	TRACK-D-5	707	0.1351	-	7471B	PPB	08 Oct 2018 14:48:44	U	SMPL
J5287-11 - 1	TRACK-D-6	210	0.0215	-	7471B	PPB	08 Oct 2018 14:51:00	U	SMPL
J5287-14 - 1	TRACK-D-7	356	0.0549	-	7471B	PPB	08 Oct 2018 14:53:16	U	SMPL
J5287-17 - 1	TRACK-D-8	1955	0.4203	-	7471B	PPB	08 Oct 2018 14:55:34	U	SMPL
J5293-01 - 1	WW-CONCRETE-WALL	9543	2.1542	-	7471B	PPB	08 Oct 2018 14:57:52	U	SMPL
J5293-03 - 1	WW-CONCRETE-FLOOR	15832	3.5912	-	7471B	PPB	08 Oct 2018 15:00:09	U	SMPL
J5294-01 - 1	HP-SUB	306	0.0435	-	7471B	PPB	08 Oct 2018 15:02:27	U	SMPL
J5300-01 - 1	SR-STONE	554	0.1001	-	7471B	PPB	08 Oct 2018 15:04:45	U	SMPL
J5305-02 - 1	331-SAND-SOIL	8422	1.8980	-	7471B	PPB	08 Oct 2018 15:07:01	U	SMPL
CCV19 - 1	CCV19	22283	5.0653	-	7471B	PPB	08 Oct 2018 15:09:18	U	SMPL
CCB19 - 1	CCB19	-124	-0.0548	-	7471B	PPB	08 Oct 2018 15:11:34	U	SMPL
J5309-01 - 1	S-9	5103	1.1396	-	7471B	PPB	08 Oct 2018 15:13:51	U	SMPL
J5309-02 - 1	S-10	2378	0.5169	-	7471B	PPB	08 Oct 2018 15:16:07	U	SMPL
J5309-03 - 1	S-11	3166	0.6970	-	7471B	PPB	08 Oct 2018 15:18:24	U	SMPL
J5309-04 - 1	S-12	3285	0.7242	-	7471B	PPB	08 Oct 2018 15:20:40	U	SMPL
J5309-05 - 1	S-13	2269	0.4920	-	7471B	PPB	08 Oct 2018 15:22:57	U	SMPL
J5309-06 - 1	S-14	4297	0.9554	-	7471B	PPB	08 Oct 2018 15:25:14	U	SMPL
J5309-07 - 1	S-15	3050	0.6705	-	7471B	PPB	08 Oct 2018 15:27:32	U	SMPL
PB113586BL - 1	PBS	7	-0.0249	-	7471B	PPB	08 Oct 2018 15:29:49	U	SMPL
PB113586BS - 1	LCSS	16102	3.6529	-	7471B	PPB	08 Oct 2018 15:32:07	U	SMPL
J5309-08 - 1	S-16	2812	0.6161	-	7471B	PPB	08 Oct 2018 15:34:24	U	SMPL
CCV20 - 1	CCV20	22265	5.0612	-	7471B	PPB	08 Oct 2018 15:36:40	U	SMPL
CCB20 - 1	CCB20	-101	-0.0495	-	7471B	PPB	08 Oct 2018 15:38:56	U	SMPL
J5309-09 - 1	S-17	1058	0.2153	-	7471B	PPB	08 Oct 2018 15:41:13	U	SMPL
J5309-10 - 1	S-18	1536	0.3245	-	7471B	PPB	08 Oct 2018 15:43:29	U	SMPL
J5309-11 - 1	S-19	1300	0.2706	-	7471B	PPB	08 Oct 2018 15:45:46	U	SMPL
J5309-12 - 1	S-20	1150	0.2363	-	7471B	PPB	08 Oct 2018 15:48:03	U	SMPL
J5309-13 - 1	S-21	3139	0.6908	-	7471B	PPB	08 Oct 2018 15:50:20	U	SMPL
J5309-14 - 1	S-22	3000	0.6590	-	7471B	PPB	08 Oct 2018 15:52:37	U	SMPL
J5309-14DUP - 1	S-22DUP	3050	0.6705	-	7471B	PPB	08 Oct 2018 15:54:54	U	SMPL
J5309-14MS - 1	S-22MS	19293	4.3821	-	7471B	PPB	08 Oct 2018 15:57:12	U	SMPL
J5309-14MSD - 1	S-22MSD	19504	4.4303	-	7471B	PPB	08 Oct 2018 15:59:30	U	SMPL
J5309-15 - 1	S-23	1716	0.3656	-	7471B	PPB	08 Oct 2018 16:01:46	U	SMPL
CCV21 - 1	CCV21	22283	5.0653	-	7471B	PPB	08 Oct 2018 16:04:03	U	SMPL
CCB21 - 1	CCB21	-174	-0.0662	-	7471B	PPB	08 Oct 2018 16:06:20	U	SMPL
J5309-16 - 1	S-24	7574	1.7042	-	7471B	PPB	08 Oct 2018 16:08:37	U	SMPL
J5309-17 - 1	S-25	8202	1.8477	-	7471B	PPB	08 Oct 2018 16:10:54	U	SMPL
J5309-18 - 1	S-26	1044	0.2121	-	7471B	PPB	08 Oct 2018 16:13:10	U	SMPL
J5309-19 - 1	S-27	1456	0.3062	-	7471B	PPB	08 Oct 2018 16:15:27	U	SMPL
J5309-20 - 1	S-28	3118	0.6860	-	7471B	PPB	08 Oct 2018 16:17:43	U	SMPL
J5309-21 - 1	S-29	1447	0.3042	-	7471B	PPB	08 Oct 2018 16:20:00	U	SMPL
J5309-22 - 1	S-30	2011	0.4331	-	7471B	PPB	08 Oct 2018 16:22:17	U	SMPL
J5309-23 - 1	S-31	4172	0.9269	-	7471B	PPB	08 Oct 2018 16:24:35	U	SMPL
J5309-24 - 1	S-32	6820	1.5319	-	7471B	PPB	08 Oct 2018 16:26:53	U	SMPL
J5309-25 - 1	S-33	2195	0.4751	-	7471B	PPB	08 Oct 2018 16:29:10	U	SMPL
CCV22 - 1	CCV22	22304	5.0701	-	7471B	PPB	08 Oct 2018 16:31:26	U	SMPL
CCB22 - 1	CCB22	-133	-0.0569	-	7471B	PPB	08 Oct 2018 16:33:43	U	SMPL
J5309-26 - 1	S-34	1767	0.3773	-	7471B	PPB	08 Oct 2018 16:36:00	U	SMPL
J5309-27 - 1	S-35	1956	0.4205	-	7471B	PPB	08 Oct 2018 16:38:17	U	SMPL
J5287-02LX5 - 1		107	-0.0020	-	7471B	PPB	08 Oct 2018 16:40:34	U	SMPL
J5287-02A - 1		18871	4.2857	-	7471B	PPB	08 Oct 2018 16:42:51	U	SMPL
J5309-14LX5 - 1		430	0.0718	-	7471B	PPB	08 Oct 2018 16:45:07	U	SMPL
J5309-14A - 1		21164	4.8096	-	7471B	PPB	08 Oct 2018 16:47:26	U	SMPL
CCV23 - 1	CCV23	22201	5.0466	-	7471B	PPB	08 Oct 2018 16:49:43	U	SMPL
CCB23 - 1	CCB23	-201	-0.0724	-	7471B	PPB	08 Oct 2018 16:52:02	U	SMPL