

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 10/08/2018

TEMP1 IN: 103 10/08/2018 07:00 TEMP OUT: 104 10/08/2018 08:10

TEMP2 IN: 104 10/08/2018 09:00 TEMP OUT: 104 10/08/2018 10:10

TEMP3 IN: 105 10/08/2018 12:30 TEMP OUT: 105 10/08/2018 14:00

TEMP4 IN: 105 10/08/2018 14:40 TEMP OUT: 105 10/08/2018 16:20

Run Number: LB98444

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB98444BL	LB98444BL	1.4125	1.4125	100	1.4125	1.4125	1.4125	0.0000	0
2	LB98444BS	LB98444BS	1.4012	1.4012	100	1.4520	1.4520	1.4520	0.0508	508
3	J5296-01	100418-JETT-E-D-B	1.3973	1.3973	250	1.4032	1.4032	1.4032	0.0059	23.6
4	J5296-01DUP	100418-JETT-E-D-BDUP	1.3814	1.3814	250	1.3876	1.3876	1.3876	0.0062	24.8
5	J5296-02	100418-JETT-E-D-S	1.3919	1.3919	500	1.3977	1.3977	1.3977	0.0058	11.6
6	J5296-03	100418-JETT-E-U-B	1.3869	1.3869	400	1.3971	1.3971	1.3971	0.0102	25.5
7	J5296-04	100418-JETT-E-U-S	1.3960	1.3960	500	1.4043	1.4043	1.4043	0.0083	16.6
8	J5296-05	100418-SIDI-E-D-B	1.3852	1.3852	500	1.3969	1.3969	1.3969	0.0117	23.4
9	J5296-06	100418-SIDI-E-D-M	1.3982	1.3982	500	1.4069	1.4069	1.4069	0.0087	17.4
10	J5296-07	100418-SIDI-E-D-S	1.4118	1.4118	500	1.4197	1.4197	1.4197	0.0079	15.8
11	J5296-08	100418-SIDI-E-U-B	1.4112	1.4112	500	1.4240	1.4240	1.4240	0.0128	25.6
12	J5296-09	100418-SIDI-E-U-M	1.3978	1.3978	300	1.4112	1.4112	1.4112	0.0134	44.7
13	J5296-10	100418-SIDI-E-U-S	1.3937	1.3937	500	1.3993	1.3993	1.3993	0.0056	11.2
14	J5296-11	100418-TIPC-E-D-M	1.4002	1.4002	500	1.4112	1.4112	1.4112	0.0110	22
15	J5296-12	100418-TIPC-E-U-M	1.4025	1.4025	500	1.4135	1.4135	1.4135	0.0110	22

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 10/08/2018

TEMP1 IN: 103 10/08/2018 07:00 TEMP OUT: 104 10/08/2018 08:10

TEMP2 IN: 104 10/08/2018 09:00 TEMP OUT: 104 10/08/2018 10:10

TEMP3 IN: 105 10/08/2018 12:30 TEMP OUT: 105 10/08/2018 14:00

TEMP4 IN: 105 10/08/2018 14:40 TEMP OUT: 105 10/08/2018 16:20

Run Number: LB98444

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J5296-13	100418-DBRI-E-D-B	1.3886	1.3886	500	1.4125	1.4125	1.4125	0.0239	47.8
17	J5296-14	100418-DBRI-E-D-S	1.3960	1.3960	400	1.4116	1.4116	1.4116	0.0156	39
18	J5296-15	100418-DBRI-E-U-B	1.3896	1.3896	400	1.4057	1.4057	1.4057	0.0161	40.3
19	J5296-16	100418-DBRI-E-U-S	1.3955	1.3955	250	1.4010	1.4010	1.4010	0.0055	22
20	J5296-16DUP	100418-DBRI-E-U-SDUP	1.3888	1.3888	250	1.3941	1.3941	1.3941	0.0053	21.2

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS100818

WorkList ID : 117666

Date : 10/8/2018 7:53:11 AM

2198444

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/11/2018	Water	J5296-01 C	TSS	Cool 4 deg C	HDR103	O31	100418-JETT-E-D-B	10/04/2018	SM2540 D
10/11/2018	Water	J5296-02 C	TSS	Cool 4 deg C	HDR103	O31	100418-JETT-E-D-S	10/04/2018	SM2540 D
10/11/2018	Water	J5296-03 C	TSS	Cool 4 deg C	HDR103	O31	100418-JETT-E-U-B	10/04/2018	SM2540 D
10/11/2018	Water	J5296-04 C	TSS	Cool 4 deg C	HDR103	O31	100418-JETT-E-U-S	10/04/2018	SM2540 D
10/11/2018	Water	J5296-05 C	TSS	Cool 4 deg C	HDR103	O31	100418-SIDI-E-D-B	10/04/2018	SM2540 D
10/11/2018	Water	J5296-06 C	TSS	Cool 4 deg C	HDR103	O31	100418-SIDI-E-D-M	10/04/2018	SM2540 D
10/11/2018	Water	J5296-07 C	TSS	Cool 4 deg C	HDR103	O31	100418-SIDI-E-D-S	10/04/2018	SM2540 D
10/11/2018	Water	J5296-08 C	TSS	Cool 4 deg C	HDR103	O31	100418-SIDI-E-U-B	10/04/2018	SM2540 D
10/11/2018	Water	J5296-09 C	TSS	Cool 4 deg C	HDR103	O31	100418-SIDI-E-U-M	10/04/2018	SM2540 D
10/11/2018	Water	J5296-10 C	TSS	Cool 4 deg C	HDR103	O31	100418-SIDI-E-U-S	10/04/2018	SM2540 D
10/09/2018	Water	J5296-11 C	TSS	Cool 4 deg C	HDR103	O31	100418-TIPC-E-D-M	10/04/2018	SM2540 D
10/09/2018	Water	J5296-12 C	TSS	Cool 4 deg C	HDR103	O31	100418-TIPC-E-U-M	10/04/2018	SM2540 D
10/11/2018	Water	J5296-13 C	TSS	Cool 4 deg C	HDR103	O31	100418-DBRI-E-D-B	10/04/2018	SM2540 D
10/11/2018	Water	J5296-14 C	TSS	Cool 4 deg C	HDR103	O31	100418-DBRI-E-D-S	10/04/2018	SM2540 D
10/11/2018	Water	J5296-15 C	TSS	Cool 4 deg C	HDR103	O31	100418-DBRI-E-U-B	10/04/2018	SM2540 D
10/11/2018	Water	J5296-16 C	TSS	Cool 4 deg C	HDR103	O31	100418-DBRI-E-U-S	10/04/2018	SM2540 D

FR- 9823888

Date/Time 10/08/18 11:40
 Received by: AS
 Relinquished by: AS

Date/Time 10/08/18 13:00
 Received by: AS
 Relinquished by: AS