



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB98450

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.9	4.01	10/08/2018	17:15
2	CAL2	1	Water			23.0	7.00	10/08/2018	17:17
3	CAL3	1	Water			22.8	10.03	10/08/2018	17:19
4	ICV	1	Water			22.9	7.01	10/08/2018	17:21
5	CCV1	1	Water			22.7	2.00	10/08/2018	17:23
6	J5200-06	1	Solid	20.02	20	23.6	8.33	10/08/2018	17:25
7	J5200-06DUP	1	Solid	20.03	20	23.7	8.34	10/08/2018	17:27
8	J5200-08	1	Solid	20.02	20	23.5	7.98	10/08/2018	17:29
9	J5201-02	1	Solid	20.02	20	23.4	7.06	10/08/2018	17:31
10	J5201-04	1	Solid	20.03	20	23.3	6.65	10/08/2018	17:33
11	J5201-06	1	Solid	20.04	20	23.3	8.60	10/08/2018	17:35
12	J5324-02	1	Solid	20.02	20	23.1	6.55	10/08/2018	17:37
13	J5325-02	1	Solid	20.02	20	23.6	7.18	10/08/2018	17:39
14	J5329-04	1	Solid	20.04	20	23.5	7.72	10/08/2018	17:41
15	J5329-08	1	Solid	20.02	20	23.5	8.28	10/08/2018	17:43
16	CCV2	1	Water			22.8	12.00	10/08/2018	17:45

WORKLIST(Hardcopy Internal Chain)

LB98450

WorkList Name : corrosion100818

WorkList ID : 117726

Date : 10/8/2018 4:55:00 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/08/2018	Solid	J5200-06	Corrosivity	Cool 4 deg C	PSEG05		OR-03-100818-A	10/08/2018	9045D
10/08/2018	Solid	J5200-08	Corrosivity	Cool 4 deg C	PSEG05		OR-03-100818-B	10/08/2018	9045D
10/08/2018	Solid	J5201-02	Corrosivity	Cool 4 deg C	PSEG05	P21	EO-02-100818-A	10/08/2018	9045D
10/08/2018	Solid	J5201-04	Corrosivity	Cool 4 deg C	PSEG05	P21	EO-02-100818-B	10/08/2018	9045D
10/08/2018	Solid	J5201-06	Corrosivity	Cool 4 deg C	PSEG05	P21	EO-02-100818-C	10/08/2018	9045D
10/08/2018	Solid	J5324-02	Corrosivity	Cool 4 deg C	PSEG01	P13	OILY-DEBRIS	10/08/2018	9045D
10/08/2018	Solid	J5325-02	Corrosivity	Cool 4 deg C	PSEG01	P13	SOIL-DEBRI	10/08/2018	9045D
10/08/2018	Solid	J5329-04	Corrosivity	Cool 4 deg C	PSEG01	P13	POND-1	10/08/2018	9045D
10/08/2018	Solid	J5329-08	Corrosivity	Cool 4 deg C	PSEG01	P13	POND-2	10/08/2018	9045D

Date/Time 10/08/18 5:00
 Received by: SP
 Relinquished by: CP

Date/Time 10/08/18 5:50
 Received by: CP
 Relinquished by: SP