



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB98452

BalanceID: WC SC-4

Reagent/Standard	Lot/Log #
COD calibration std. 50 ppm	WP67995
COD calibration std. 0 ppm	WP67992
COD Digestion Vials Low Level 0-150Mg/L	W2395
COD ICV-LCS std, 50ppm	WP67997
COD calibration std. 100 ppm	WP67994
COD calibration std. 10 ppm	WP67993
COD calibration std. 150 ppm	WP67996
COD ICV-LCS std, 50ppm	WP68491
COD CCV std, 50ppm	WP68490
COD LOD std, 5ppm	WP68469
COD LOQ std, 10.0PPM	WP68470

Temp In (C): 150	Date In: 10/08/2018	Time In: 11:00
Temp Out (C): 150	Date Out: 10/08/2018	Time Out: 13:00

Intercept: -0.0516

Slope: 0.9976

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	09/20/2018	13:30
2	CAL2	10	1	Water	10.000	09/20/2018	13:30
3	CAL3	50	1	Water	50.000	09/20/2018	13:31
4	CAL4	100	1	Water	99.000	09/20/2018	13:31
5	CAL5	150	1	Water	150.000	09/20/2018	13:32



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB98452

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	49.000	49.170	09/20/2018	13:32
2	ICB				1	Water	0.000	0.052	09/20/2018	13:33
3	CCV1	50			1	Water	50.000	50.172	10/08/2018	14:10
4	CCB1				1	Water	0.000	0.052	10/08/2018	14:10
5	LB98452BL		1.00	50	1	Solid	0.000	0.052	10/08/2018	14:11
6	LB98452BS	50	1.01	50	1	Solid	50.000	50.172	10/08/2018	14:11
7	J5164-01		1.00	50	1	Solid	5.000	5.064	10/08/2018	14:12
8	J5164-02		1.02	50	1	Solid	11.000	11.078	10/08/2018	14:12
9	J5164-03		1.00	50	1	Solid	6.000	6.066	10/08/2018	14:13
10	CCV2	50			1	Water	51.000	51.174	10/08/2018	14:13
11	CCB2				1	Water	0.000	0.052	10/08/2018	14:14