

LB 98456


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

35482



A= 0.0000e+000

B= 2.7974e-004

C= 2.5112e-002

Rho= 0.9998004

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	-39	0.000	-39				
0.2	0.200	0.216	0.016	684	0.0 %	684				
2.5	2.500	2.377	-0.123	8408	0.0 %	8408				
5.0	5.000	5.119	0.119	18209	0.0 %	18209				
7.5	7.500	7.522	0.022	26800	0.0 %	26800				
10.0	10.000	9.951	-0.049	35482	0.0 %	35482				

LB98456 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Oct 2018 07:51:41

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	SO	-39	-	0.00007470A	PPB	09 Oct 2018 08:32:46	S	Std
0.2 - 1	SO.2	684	-	0.20007470A	PPB	09 Oct 2018 08:34:55	S	Std
2.5 - 1	SO.5	8408	-	2.50007470A	PPB	09 Oct 2018 08:37:03	S	Std
5.0 - 1	SO.5	18209	-	5.00007470A	PPB	09 Oct 2018 08:39:12	S	Std
7.5 - 1	SO.5	26800	-	7.50007470A	PPB	09 Oct 2018 08:41:20	S	Std
10.0 - 1	SO.5	35482	-	10.00007470A	PPB	09 Oct 2018 08:45:57	S	Std
ICV12 - 1	ICV12	14751	4.1516	-7470A	PPB	09 Oct 2018 08:48:25	U	SMPL
ICB12 - 1	ICB12	-239	-0.0417	-7470A	PPB	09 Oct 2018 08:50:33	U	SMPL
CCV24 - 1	CCV24	18817	5.2890	-7470A	PPB	09 Oct 2018 08:52:41	U	SMPL
CCB24 - 1	CCB24	-316	-0.0633	-7470A	PPB	09 Oct 2018 08:54:48	U	SMPL
CRA - 1	CRA	620	0.1986	-7470A	PPB	09 Oct 2018 08:56:57	U	SMPL
HighStd - 1	HighStd	35594	9.9823	-7470A	PPB	09 Oct 2018 08:59:03	U	SMPL
ChkStd - 1	ChkStd	26012	7.3018	-7470A	PPB	09 Oct 2018 09:01:10	U	SMPL
PB113625BL - 1	PBW	-333	-0.0680	-7470A	PPB	09 Oct 2018 09:03:19	U	SMPL
PB113625BS - 1	LCSW	12121	3.4159	-7470A	PPB	09 Oct 2018 09:05:28	U	SMPL
PB113625TB - 1	PB112583TB	-247	-0.0440	-7470A	PPB	09 Oct 2018 09:07:36	U	SMPL
J5270-08 - 1	PHASE-3-SOIL	-93	-0.0009	-7470A	PPB	09 Oct 2018 09:09:46	U	SMPL
J5270-08DUP - 1	PHASE-3-SOILDUP	-70	0.0055	-7470A	PPB	09 Oct 2018 09:11:54	U	SMPL
J5270-08MS - 1	PHASE-3-SOILMS	11872	3.3462	-7470A	PPB	09 Oct 2018 09:14:02	U	SMPL
J5270-08MSD - 1	PHASE-3-SOILMSD	11935	3.3638	-7470A	PPB	09 Oct 2018 09:18:35	U	SMPL
CCV25 - 1	CCV25	18641	5.2398	-7470A	PPB	09 Oct 2018 09:20:41	U	SMPL
CCB25 - 1	CCB25	-271	-0.0507	-7470A	PPB	09 Oct 2018 09:22:49	U	SMPL
J5287-03 - 1	TRACK-D-3	-42	0.0134	-7470A	PPB	09 Oct 2018 09:24:56	U	SMPL
J5287-06 - 1	TRACK-D-4	-35	0.0153	-7470A	PPB	09 Oct 2018 09:27:04	U	SMPL
J5287-09 - 1	TRACK-D-5	-23	0.0187	-7470A	PPB	09 Oct 2018 09:29:10	U	SMPL
J5287-12 - 1	TRACK-D-6	-49	0.0114	-7470A	PPB	09 Oct 2018 09:31:17	U	SMPL
J5287-15 - 1	TRACK-D-7	-99	-0.0026	-7470A	PPB	09 Oct 2018 09:33:25	U	SMPL
J5287-18 - 1	TRACK-D-8	-86	0.0011	-7470A	PPB	09 Oct 2018 09:35:32	U	SMPL
J5293-02 - 1	VW-CONCRETE-WALL	0	0.0251	-7470A	PPB	09 Oct 2018 09:37:40	U	SMPL
J5293-04 - 1	VW-CONCRETE-FLOOR	24	0.0318	-7470A	PPB	09 Oct 2018 09:39:49	U	SMPL
J5294-02 - 1	HP-SUB	-8	0.0229	-7470A	PPB	09 Oct 2018 09:41:58	U	SMPL
J5300-04 - 1	SR-STONE	-60	0.0083	-7470A	PPB	09 Oct 2018 09:44:05	U	SMPL
CCV26 - 1	CCV26	18579	5.2225	-7470A	PPB	09 Oct 2018 09:46:11	U	SMPL
CCB26 - 1	CCB26	-223	-0.0373	-7470A	PPB	09 Oct 2018 09:48:18	U	SMPL
J5305-03 - 1	331-SAND-SOIL	-32	0.0162	-7470A	PPB	09 Oct 2018 09:50:26	U	SMPL
J5307-01 - 1	148-20-OLD	28	0.0329	-7470A	PPB	09 Oct 2018 09:52:33	U	SMPL
J5270-08LX5 - 1		-78	0.0033	-7470A	PPB	09 Oct 2018 09:54:40	U	SMPL
J5270-08A - 1		13618	3.8347	-7470A	PPB	09 Oct 2018 09:56:47	U	SMPL
CCV27 - 1	CCV27	18714	5.2602	-7470A	PPB	09 Oct 2018 09:58:55	U	SMPL
CCB27 - 1	CCB27	-198	-0.0303	-7470A	PPB	09 Oct 2018 10:01:02	U	SMPL