



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB98458

Slope : 97.3

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.7	4.01	10/09/2018	09:45
2	CAL2	1	Water			22.8	7.00	10/09/2018	09:47
3	CAL3	1	Water			22.7	10.03	10/09/2018	09:49
4	ICV	1	Water			22.9	7.01	10/09/2018	09:51
5	CCV1	1	Water			22.8	2.00	10/09/2018	09:53
6	J5326-04	1	Solid	20.03	20	22.9	6.05	10/09/2018	09:55
7	J5326-04DUP	1	Solid	20.04	20	23.0	6.06	10/09/2018	09:57
8	J5326-08	1	Solid	20.02	20	22.8	5.82	10/09/2018	09:59
9	J5326-12	1	Solid	20.02	20	22.8	5.98	10/09/2018	10:01
10	J5326-16	1	Solid	20.03	20	22.7	6.93	10/09/2018	10:03
11	J5326-20	1	Solid	20.02	20	22.7	6.38	10/09/2018	10:05
12	J5326-24	1	Solid	20.04	20	22.5	5.31	10/09/2018	10:07
13	J5326-28	1	Solid	20.02	20	22.7	5.69	10/09/2018	10:09
14	J5326-32	1	Solid	20.03	20	22.6	5.60	10/09/2018	10:11
15	J5326-35	1	Solid	20.02	20	22.6	5.76	10/09/2018	10:13
16	CCV2	1	Water			22.7	12.00	10/09/2018	10:15
17	J5326-38	1	Solid	20.02	20	22.7	6.32	10/09/2018	10:17
18	J5326-41	1	Solid	20.03	20	22.6	6.07	10/09/2018	10:19
19	J5331-03	1	Solid	20.03	20	22.3	5.43	10/09/2018	10:21
20	J5331-03DUP	1	Solid	20.03	20	22.4	5.44	10/09/2018	10:23

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water			22.6	2.00	10/09/2018	10:25

WORKLIST(Hardcopy Internal Chain)

LB98u58

WorkList Name : CORROSIONITY100918

WorkList ID : 117758

Date : 10/9/2018 9:23:07 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/08/2018	Solid	J5326-04	Corrosivity	Cool 4 deg C	PSEG01	P21	FL-PIT-1	10/08/2018	9045D
10/08/2018	Solid	J5326-08	Corrosivity	Cool 4 deg C	PSEG01	P21	FL-PIT-2	10/08/2018	9045D
10/08/2018	Solid	J5326-12	Corrosivity	Cool 4 deg C	PSEG01	P21	FL-PIT-3	10/08/2018	9045D
10/08/2018	Solid	J5326-16	Corrosivity	Cool 4 deg C	PSEG01	P21	FL-PIT-4	10/08/2018	9045D
10/08/2018	Solid	J5326-20	Corrosivity	Cool 4 deg C	PSEG01	P21	FL-PIT-5	10/08/2018	9045D
10/08/2018	Solid	J5326-24	Corrosivity	Cool 4 deg C	PSEG01	P21	FL-PIT-6	10/08/2018	9045D
10/08/2018	Solid	J5326-28	Corrosivity	Cool 4 deg C	PSEG01	P21	FL-PIT-7	10/08/2018	9045D
10/08/2018	Solid	J5326-32	Corrosivity	Cool 4 deg C	PSEG01	P21	FL-PIT-8	10/08/2018	9045D
10/08/2018	Solid	J5326-35	Corrosivity	Cool 4 deg C	PSEG01	P21	FL-PH-2-STONE	10/08/2018	9045D
10/08/2018	Solid	J5326-38	Corrosivity	Cool 4 deg C	PSEG01	P21	FL-PH-1-STONE	10/08/2018	9045D
10/08/2018	Solid	J5326-41	Corrosivity	Cool 4 deg C	PSEG01	P21	FL-PH-3-STONE	10/08/2018	9045D
10/08/2018	Solid	J5331-03	Corrosivity	Cool 4 deg C	PSEG01	P21	NM-STONE-DRUM-3	10/08/2018	9045D

Date/Time 10/09/18 9:25

Received by: SP

Relinquished by: CP

Date/Time 10/09/18 10:35

Received by: CP

Relinquished by: SP