

LB98459


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

34595



A= 0.0000e+000

B= 2.8423e-004

C= -3.9247e-002

Rho= 0.9995054

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.054	-0.054	-52	0.000	-52				
0.05	0.050	0.054	0.004	327	0.0 %	327				
0.2	0.200	0.143	-0.057	641	0.0 %	641				
2.5	2.500	2.507	0.007	8959	0.0 %	8959				
5.0	5.000	5.136	0.136	18209	0.0 %	18209				
7.5	7.500	7.669	0.169	27119	0.0 %	27119				
10.0	10.000	9.795	-0.205	34599	0.0 %	34599				

LB98459 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 09 Oct 2018 10:28:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-52	-	0.0000	245.1	PPB	09 Oct 2018 10:32:08	S	Std
0.05 - 1	50.05	327	-	0.0500	245.1	PPB	09 Oct 2018 10:37:23	S	Std
0.2 - 1	50.2	641	-	0.2000	245.1	PPB	09 Oct 2018 10:39:41	S	Std
2.5 - 1	52.5	8959	-	2.5000	245.1	PPB	09 Oct 2018 10:41:58	S	Std
5.0 - 1	55	18209	-	5.0000	245.1	PPB	09 Oct 2018 10:44:15	S	Std
7.5 - 1	57.5	27119	-	7.5000	245.1	PPB	09 Oct 2018 10:46:31	S	Std
10.0 - 1	510	34599	-	10.0000	245.1	PPB	09 Oct 2018 10:51:34	S	Std
ICV13 - 1	ICV13	14246	4.0099	-245.1		PPB	09 Oct 2018 10:55:01	U	SMPL
ICB13 - 1	ICB13	-10	-0.0421	-245.1		PPB	09 Oct 2018 10:59:51	U	SMPL
CCV28 - 1	CCV28	17869	5.0397	-245.1		PPB	09 Oct 2018 11:02:07	U	SMPL
CCB28 - 1	CCB28	-67	-0.0583	-245.1		PPB	09 Oct 2018 11:07:00	U	SMPL
CRA - 1	CRA	304	0.0472	-245.1		PPB	09 Oct 2018 11:14:29	U	SMPL
HighStd - 1	HighStd	35429	10.0309	-245.1		PPB	09 Oct 2018 11:16:44	U	SMPL
ChkStd - 1	ChkStd	24661	6.9702	-245.1		PPB	09 Oct 2018 11:19:01	U	SMPL
PB113641BL - 1	PBW	-15	-0.0435	-245.1		PPB	09 Oct 2018 11:24:26	U	SMPL
PB113641BS - 1	LCSW	12899	3.6271	-245.1		PPB	09 Oct 2018 11:26:42	U	SMPL
J5264-01 - 1	100318-SIDI-E-D-B	-89	-0.0645	-245.1		PPB	09 Oct 2018 11:29:00	U	SMPL
J5264-02 - 1	100318-SIDI-E-D-M	-54	-0.0546	-245.1		PPB	09 Oct 2018 11:31:17	U	SMPL
J5264-03 - 1	100318-SIDI-E-D-S	-31	-0.0481	-245.1		PPB	09 Oct 2018 11:33:35	U	SMPL
J5264-04 - 1	100318-SIDI-E-U-B	-19	-0.0446	-245.1		PPB	09 Oct 2018 11:35:53	U	SMPL
J5264-05 - 1	100318-SIDI-E-U-M	79	-0.0168	-245.1		PPB	09 Oct 2018 11:38:08	U	SMPL
CCV29 - 1	CCV29	17882	5.0434	-245.1		PPB	09 Oct 2018 11:40:24	U	SMPL
CCB29 - 1	CCB29	-178	-0.0898	-245.1		PPB	09 Oct 2018 11:42:40	U	SMPL
J5264-06 - 1	100318-SIDI-E-U-S	-87	-0.0640	-245.1		PPB	09 Oct 2018 11:44:57	U	SMPL
J5264-07 - 1	100318-TIPC-E-D-M	-7	-0.0412	-245.1		PPB	09 Oct 2018 11:47:13	U	SMPL
J5264-08 - 1	100318-TIPC-E-U-M	-131	-0.0765	-245.1		PPB	09 Oct 2018 11:49:30	U	SMPL
J5264-09 - 1	100318-DBRI-E-D-B	49	-0.0253	-245.1		PPB	09 Oct 2018 11:51:46	U	SMPL
J5264-10 - 1	100318-DBRI-E-D-S	-54	-0.0546	-245.1		PPB	09 Oct 2018 11:54:03	U	SMPL
J5264-11 - 1	100318-DBRI-E-U-B	-78	-0.0614	-245.1		PPB	09 Oct 2018 11:56:20	U	SMPL
J5264-12 - 1	100318-DBRI-E-U-S	-72	-0.0597	-245.1		PPB	09 Oct 2018 11:58:37	U	SMPL
J5264-12DUP - 1	100318-DBRI-E-U-SDUP	-81	-0.0623	-245.1		PPB	09 Oct 2018 12:00:55	U	SMPL
J5264-12MS - 1	100318-DBRI-E-U-SMS	14658	4.1270	-245.1		PPB	09 Oct 2018 12:03:13	U	SMPL
J5264-12MSD - 1	100318-DBRI-E-U-SMSD	14893	4.1938	-245.1		PPB	09 Oct 2018 12:05:29	U	SMPL
CCV30 - 1	CCV30	17461	4.9238	-245.1		PPB	09 Oct 2018 12:07:46	U	SMPL
CCB30 - 1	CCB30	-192	-0.0938	-245.1		PPB	09 Oct 2018 12:12:58	U	SMPL
PB113642BL - 1	PBW	-26	-0.0466	-245.1		PPB	09 Oct 2018 12:15:14	U	SMPL
PB113642BS - 1	LCSW	12884	3.6228	-245.1		PPB	09 Oct 2018 12:17:31	U	SMPL
J5296-01 - 1	100418-JETT-E-D-B	-76	-0.0608	-245.1		PPB	09 Oct 2018 12:19:47	U	SMPL
J5296-02 - 1	100418-JETT-E-D-S	124	-0.0040	-245.1		PPB	09 Oct 2018 12:22:06	U	SMPL
J5296-03 - 1	100418-JETT-E-U-B	27	-0.0316	-245.1		PPB	09 Oct 2018 12:24:23	U	SMPL
J5296-04 - 1	100418-JETT-E-U-S	23	-0.0327	-245.1		PPB	09 Oct 2018 12:26:40	U	SMPL
J5296-05 - 1	100418-SIDI-E-D-B	4	-0.0381	-245.1		PPB	09 Oct 2018 12:28:58	U	SMPL
J5296-06 - 1	100418-SIDI-E-D-M	-96	-0.0665	-245.1		PPB	09 Oct 2018 12:31:16	U	SMPL
J5296-07 - 1	100418-SIDI-E-D-S	-13	-0.0429	-245.1		PPB	09 Oct 2018 12:33:33	U	SMPL
J5296-08 - 1	100418-SIDI-E-U-B	-51	-0.0537	-245.1		PPB	09 Oct 2018 12:35:50	U	SMPL
CCV31 - 1	CCV31	17598	4.9627	-245.1		PPB	09 Oct 2018 12:42:57	U	SMPL
CCB31 - 1	CCB31	-283	-0.1197	-245.1		PPB	09 Oct 2018 12:45:13	U	SMPL
J5296-09 - 1	100418-SIDI-E-U-M	-60	-0.0563	-245.1		PPB	09 Oct 2018 12:47:31	U	SMPL
J5296-10 - 1	100418-SIDI-E-U-S	-74	-0.0603	-245.1		PPB	09 Oct 2018 12:49:47	U	SMPL
J5296-11 - 1	100418-TIPC-E-D-M	-66	-0.0580	-245.1		PPB	09 Oct 2018 12:52:03	U	SMPL
J5296-12 - 1	100418-TIPC-E-U-M	-31	-0.0481	-245.1		PPB	09 Oct 2018 12:54:20	U	SMPL
J5296-13 - 1	100418-DBRI-E-D-B	-65	-0.0577	-245.1		PPB	09 Oct 2018 12:56:36	U	SMPL
J5296-14 - 1	100418-DBRI-E-D-S	-3	-0.0401	-245.1		PPB	09 Oct 2018 12:58:54	U	SMPL
J5296-15 - 1	100418-DBRI-E-U-B	-25	-0.0464	-245.1		PPB	09 Oct 2018 13:01:11	U	SMPL
J5296-16 - 1	100418-DBRI-E-U-S	-67	-0.0583	-245.1		PPB	09 Oct 2018 13:03:29	U	SMPL
J5296-16DUP - 1	100418-DBRI-E-U-SDUP	-30	-0.0478	-245.1		PPB	09 Oct 2018 13:05:46	U	SMPL
J5296-16MS - 1	100418-DBRI-E-U-SMS	13478	3.7917	-245.1		PPB	09 Oct 2018 13:08:03	U	SMPL
CCV32 - 1	CCV32	17183	4.8447	-245.1		PPB	09 Oct 2018 13:10:19	U	SMPL
CCB32 - 1	CCB32	-117	-0.0725	-245.1		PPB	09 Oct 2018 13:12:35	U	SMPL
J5296-16MSD - 1	100418-DBRI-E-U-SMSD	13745	3.8675	-245.1		PPB	09 Oct 2018 13:14:54	U	SMPL
PB113643BL - 1	PBW	-222	-0.1023	-245.1		PPB	09 Oct 2018 13:17:10	U	SMPL
PB113643BS - 1	LCSW	12610	3.5449	-245.1		PPB	09 Oct 2018 13:19:26	U	SMPL
J5227-02 - 1	COMP	146	0.0023	-245.1		PPB	09 Oct 2018 13:21:43	U	SMPL
J5227-02DUP - 1	COMPDUP	101	-0.0105	-245.1		PPB	09 Oct 2018 13:23:59	U	SMPL
J5227-02MS - 1	COMPMS	12555	3.5293	-245.1		PPB	09 Oct 2018 13:26:17	U	SMPL
J5227-02MSD - 1	COMPMSD	12742	3.5825	-245.1		PPB	09 Oct 2018 13:28:34	U	SMPL
J5231-02 - 1	EFF-WW	-74	-0.0603	-245.1		PPB	09 Oct 2018 13:30:53	U	SMPL
J5299-01 - 1	100518-DBRI-F-D-B	-17	-0.0441	-245.1		PPB	09 Oct 2018 13:33:13	U	SMPL
J5299-02 - 1	100518-DBRI-F-D-S	-50	-0.0535	-245.1		PPB	09 Oct 2018 13:35:29	U	SMPL
CCV33 - 1	CCV33	18057	5.0932	-245.1		PPB	09 Oct 2018 13:37:46	U	SMPL
CCB33 - 1	CCB33	-280	-0.1188	-245.1		PPB	09 Oct 2018 13:40:03	U	SMPL
J5299-03 - 1	100518-DBRI-F-U-B	-93	-0.0657	-245.1		PPB	09 Oct 2018 13:42:21	U	SMPL
J5299-04 - 1	100518-DBRI-F-U-S	3	-0.0384	-245.1		PPB	09 Oct 2018 13:44:38	U	SMPL
J5299-05 - 1	100518-SIDI-F-D-B	-20	-0.0449	-245.1		PPB	09 Oct 2018 13:46:54	U	SMPL
J5299-06 - 1	100518-SIDI-F-D-M	-35	-0.0492	-245.1		PPB	09 Oct 2018 13:49:11	U	SMPL
J5299-07 - 1	100518-SIDI-F-D-S	4	-0.0381	-245.1		PPB	09 Oct 2018 13:51:27	U	SMPL
J5299-08 - 1	100518-SIDI-F-U-B	-54	-0.0546	-245.1		PPB	09 Oct 2018 13:53:45	U	SMPL
J5299-09 - 1	100518-SIDI-F-U-M	109	-0.0083	-245.1		PPB	09 Oct 2018 13:56:02	U	SMPL
J5299-10 - 1	100518-SIDI-F-U-S	47	-0.0259	-245.1		PPB	09 Oct 2018 13:58:20	U	SMPL
CCV34 - 1	CCV34	18262	5.1514	-245.1		PPB	09 Oct 2018 14:00:38	U	SMPL
CCB34 - 1	CCB34	-240	-0.1075	-245.1		PPB	09 Oct 2018 14:02:54	U	SMPL

LB98459 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 09 Oct 2018 10:28:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
J5264-12LX5 - 1		194	0.0159	-245.1	PPB		09 Oct 2018 14:05:13	U	SMPL
J5264-12A - 1		14571	4.1023	-245.1	PPB		09 Oct 2018 14:07:31	U	SMPL
J5296-16LX5 - 1		104	-0.0097	-245.1	PPB		09 Oct 2018 14:09:48	U	SMPL
J5296-16A - 1		14403	4.0546	-245.1	PPB		09 Oct 2018 14:12:07	U	SMPL
J5227-02LX5 - 1		114	-0.0068	-245.1	PPB		09 Oct 2018 14:14:24	U	SMPL
J5227-02A - 1		14256	4.0128	-245.1	PPB		09 Oct 2018 14:16:42	U	SMPL
CCV35 - 1	CCV35	18404	5.1918	-245.1	PPB		09 Oct 2018 14:21:02	U	SMPL
CCB35 - 1	CCB35	-261	-0.1134	-245.1	PPB		09 Oct 2018 14:23:19	U	SMPL